

LB115751


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

52196



A= 0.0000e+000

B= 1.9095e-004

C= -2.9570e-002

Rho= 0.9998943

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.026	0.026	293	0.000	293					
0.2	0.200	0.187	-0.013	1136	0.0 %	1136					
2.5	2.500	2.431	-0.069	12887	0.0 %	12887					
5.0	5.000	5.039	0.039	26547	0.0 %	26547					
7.5	7.500	7.579	0.079	39844	0.0 %	39844					
10.0	10.000	9.937	-0.063	52196	0.0 %	52196					

LB115751 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Aug 2021 12:03:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		293	-	0.0000	7471B	PPB	09 Aug 2021 12:04:50	S	Std
0.2 - 1		1136	-	0.2000	7471B	PPB	09 Aug 2021 12:07:06	S	Std
2.5 - 1		12887	-	2.5000	7471B	PPB	09 Aug 2021 12:09:23	S	Std
5.0 - 1		26547	-	5.0000	7471B	PPB	09 Aug 2021 12:11:40	S	Std
7.5 - 1		39844	-	7.5000	7471B	PPB	09 Aug 2021 12:13:57	S	Std
10.0 - 1		52196	-	10.0000	7471B	PPB	09 Aug 2021 12:16:12	S	Std
ICV44 - 1	ICV44	20730	3.8264	-	7471B	PPB	09 Aug 2021 12:20:24	U	SMPL
ICB44 - 1	ICB44	18	-0.0168	-	7471B	PPB	09 Aug 2021 12:22:38	U	SMPL
CCV13 - 1	CCV13	25367	4.6868	-	7471B	PPB	09 Aug 2021 12:24:53	U	SMPL
CCB13 - 1	CCB13	-10	-0.0220	-	7471B	PPB	09 Aug 2021 12:27:08	U	SMPL
CRA - 1	CRA	1143	0.1919	-	7471B	PPB	09 Aug 2021 12:29:24	U	SMPL
HighStd - 1	HighStd	52651	9.7495	-	7471B	PPB	09 Aug 2021 12:31:39	U	SMPL
ChkStd - 1	ChkStd	36930	6.8324	-	7471B	PPB	09 Aug 2021 12:33:54	U	SMPL
PB138256BL - 1	PBS	-120	-0.0424	-	7471B	PPB	09 Aug 2021 12:36:12	U	SMPL
PB138256BS - 1	LCSS	19531	3.6039	-	7471B	PPB	09 Aug 2021 12:38:28	U	SMPL
M3280-01 - 1	TR-04-080421	1183	0.1993	-	7471B	PPB	09 Aug 2021 12:40:45	U	SMPL
M3280-01DUP - 1	TR-04-080421DUP	1078	0.1799	-	7471B	PPB	09 Aug 2021 12:43:01	U	SMPL
M3280-01MS - 1	TR-04-080421MS	20400	3.7652	-	7471B	PPB	09 Aug 2021 12:45:18	U	SMPL
M3280-01MSD - 1	TR-04-080421MSD	19569	3.6110	-	7471B	PPB	09 Aug 2021 12:47:36	U	SMPL
M3281-01 - 1	TWP03	2525	0.4484	-	7471B	PPB	09 Aug 2021 12:49:51	U	SMPL
CCV14 - 1	CCV14	25388	4.6907	-	7471B	PPB	09 Aug 2021 12:52:07	U	SMPL
CCB14 - 1	CCB14	11	-0.0181	-	7471B	PPB	09 Aug 2021 12:54:24	U	SMPL
M3289-01 - 1	27193	6080	1.1080	-	7471B	PPB	09 Aug 2021 12:56:40	U	SMPL
M3289-03 - 1	26240	1009	0.1671	-	7471B	PPB	09 Aug 2021 12:58:55	U	SMPL
M3289-05 - 1	44043RT	3852	0.6946	-	7471B	PPB	09 Aug 2021 13:01:10	U	SMPL
M3289-07 - 1	26562	2719	0.4844	-	7471B	PPB	09 Aug 2021 13:03:25	U	SMPL
M3289-09 - 1	37702RT	156	0.0088	-	7471B	PPB	09 Aug 2021 13:05:41	U	SMPL
M3289-11 - 1	26498	13820	2.5442	-	7471B	PPB	09 Aug 2021 13:07:57	U	SMPL
M3289-13 - 1	25082	10117	1.8571	-	7471B	PPB	09 Aug 2021 13:10:14	U	SMPL
M3291-01 - 1	WC-1	6137	1.1186	-	7471B	PPB	09 Aug 2021 13:12:31	U	SMPL
M3291-04 - 1	WC-2	365	0.0476	-	7471B	PPB	09 Aug 2021 13:14:49	U	SMPL
M3292-01 - 1	25546	4680	0.8482	-	7471B	PPB	09 Aug 2021 13:17:04	U	SMPL
CCV15 - 1	CCV15	25259	4.6668	-	7471B	PPB	09 Aug 2021 13:19:20	U	SMPL
CCB15 - 1	CCB15	17	-0.0170	-	7471B	PPB	09 Aug 2021 13:21:35	U	SMPL
M3293-01 - 1	289	2600	0.4623	-	7471B	PPB	09 Aug 2021 13:23:51	U	SMPL
M3293-03 - 1	20071	656	0.1016	-	7471B	PPB	09 Aug 2021 13:26:07	U	SMPL
M3293-05 - 1	1927	83	-0.0048	-	7471B	PPB	09 Aug 2021 13:28:22	U	SMPL
M3293-07 - 1	2071	3881	0.7000	-	7471B	PPB	09 Aug 2021 13:30:37	U	SMPL
M3294-01 - 1	RBR-018	2822	0.5035	-	7471B	PPB	09 Aug 2021 13:32:54	U	SMPL
M3295-01 - 1	FK-EF-01-022-ABCD	1409	0.2413	-	7471B	PPB	09 Aug 2021 13:35:10	U	SMPL
M3296-01 - 1	FK-BPM-03-026-ABCD	2253	0.3979	-	7471B	PPB	09 Aug 2021 13:37:26	U	SMPL
M3297-01 - 1	128	185	0.0142	-	7471B	PPB	09 Aug 2021 13:39:43	U	SMPL
PB138257BL - 1	PBS	57	-0.0096	-	7471B	PPB	09 Aug 2021 13:42:00	U	SMPL
PB138257BS - 1	LCSS	19265	3.5545	-	7471B	PPB	09 Aug 2021 13:44:16	U	SMPL
CCV16 - 1	CCV16	24814	4.5842	-	7471B	PPB	09 Aug 2021 13:46:31	U	SMPL
CCB16 - 1	CCB16	133	0.0045	-	7471B	PPB	09 Aug 2021 13:48:46	U	SMPL
M3279-01 - 1	20210568-COM-41	29638	5.4793	-	7471B	PPB	09 Aug 2021 13:51:01	U	SMPL
M3279-02 - 1	20210569-COM-42	24047	4.4419	-	7471B	PPB	09 Aug 2021 13:53:17	U	SMPL
M3279-03 - 1	20210570-COM-43	39111	7.2371	-	7471B	PPB	09 Aug 2021 13:55:33	U	SMPL
M3279-04 - 1	20210571-COM-44	29967	5.5404	-	7471B	PPB	09 Aug 2021 13:57:49	U	SMPL
M3279-05 - 1	20210572-COM-45	65773	12.1843	-	7471B	PPB	09 Aug 2021 14:00:06	U	SMPL
M3279-06 - 1	20210573-COM-46	80134	14.8491	-	7471B	PPB	09 Aug 2021 14:02:23	U	SMPL
M3279-07 - 1	20210577-COM-50	58667	10.8658	-	7471B	PPB	09 Aug 2021 14:04:41	U	SMPL
M3279-08 - 1	20210580-COM-3	48659	9.0087	-	7471B	PPB	09 Aug 2021 14:07:00	U	SMPL
M3279-09 - 1	20210589-AMENDED-COM-3	41052	7.5972	-	7471B	PPB	09 Aug 2021 14:09:18	U	SMPL
CCV17 - 1	CCV17	25265	4.6679	-	7471B	PPB	09 Aug 2021 14:11:36	U	SMPL
CCB17 - 1	CCB17	132	0.0043	-	7471B	PPB	09 Aug 2021 14:13:53	U	SMPL
M3279-11 - 1	20210532-COM-5	11028	2.0261	-	7471B	PPB	09 Aug 2021 14:16:09	U	SMPL
M3279-11DUP - 1	20210532-COM-5DUP	11176	2.0536	-	7471B	PPB	09 Aug 2021 14:18:25	U	SMPL
M3279-11MS - 1	20210532-COM-5MS	31052	5.7417	-	7471B	PPB	09 Aug 2021 14:20:40	U	SMPL
M3279-11MSD - 1	20210532-COM-5MSD	29186	5.3954	-	7471B	PPB	09 Aug 2021 14:22:55	U	SMPL
M3280-01LX5 - 1		46	-0.0116	-	7471B	PPB	09 Aug 2021 14:25:12	U	SMPL
M3280-01A - 1		19618	3.6200	-	7471B	PPB	09 Aug 2021 14:27:28	U	SMPL
M3279-11LX5 - 1		2015	0.3537	-	7471B	PPB	09 Aug 2021 14:29:45	U	SMPL
M3279-11A - 1		31884	5.8961	-	7471B	PPB	09 Aug 2021 14:32:01	U	SMPL
M3279-05DLX2 - 1		31967	5.9115	-	7471B	PPB	09 Aug 2021 14:34:17	U	SMPL
M3279-06DLX2 - 1		39322	7.2762	-	7471B	PPB	09 Aug 2021 14:36:34	U	SMPL
CCV18 - 1	CCV18	26311	4.8620	-	7471B	PPB	09 Aug 2021 14:50:32	U	SMPL
CCB18 - 1	CCB18	75	-0.0062	-	7471B	PPB	09 Aug 2021 14:52:48	U	SMPL
M3279-07DLX2 - 1		27693	5.1184	-	7471B	PPB	09 Aug 2021 14:55:05	U	SMPL
CCV19 - 1	CCV19	25743	4.7566	-	7471B	PPB	09 Aug 2021 14:57:21	U	SMPL
CCB19 - 1	CCB19	86	-0.0042	-	7471B	PPB	09 Aug 2021 14:59:38	U	SMPL