

LB 115752


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

48097



A= 0.0000e+000

B= 2.0988e-004 *slope*C= -1.7497e-002 *-intercept*

Rho= 0.9999079

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.029	0.029	221	0.000	221					
0.05	0.050					573					
0.20	0.200	0.207	0.007	1068	0.0 %	1068					
2.50	2.500	2.510	0.010	12044	0.0 %	12044					
5.00	5.000	4.956	-0.044	23699	0.0 %	23699					
7.50	7.500	7.421	-0.079	35442	0.0 %	35442					
10.0	10.000	10.077	0.077	48097	0.0 %	48097					

LB115752

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	221	-		0 SFAM01.1	PPB	8/9/2021 11:06	Std
	0.2 S01	1068	-		0.2 SFAM01.1	PPB	8/9/2021 11:08	Std
	2.5 S02	12044	-		2.5 SFAM01.1	PPB	8/9/2021 11:11	Std
	5 S03	23699	-		5 SFAM01.1	PPB	8/9/2021 11:13	Std
	7.5 S04	35442	-		7.5 SFAM01.1	PPB	8/9/2021 11:15	Std
	10 S05	48097	-		10 SFAM01.1	PPB	8/9/2021 11:17	Std
ICV058	ICV058	19516	4.0784 -		SFAM01.1	PPB	8/9/2021 11:20	SMPL
ICB058	ICB058	113	0.0062 -		SFAM01.1	PPB	8/9/2021 11:22	SMPL
CCV019	CCV019	23790	4.9755 -		SFAM01.1	PPB	8/9/2021 11:25	SMPL
CCB019	CCB019	93	0.002 -		SFAM01.1	PPB	8/9/2021 11:27	SMPL
PB138278BL	PBS278	48	-0.0074 -		SFAM01.1	PPB	8/9/2021 11:29	SMPL
M3213-04	MH0AD2	207482	43.528 -		SFAM01.1	PPB	8/9/2021 11:31	SMPL
M3213-04DLX10	MH0AD2	22349	4.673 -		SFAM01.1	PPB	8/9/2021 11:36	SMPL
CCV020	CCV020	23160	4.8432 -		SFAM01.1	PPB	8/9/2021 11:38	SMPL
CCB020	CCB020	97	0.0029 -		SFAM01.1	PPB	8/9/2021 11:40	SMPL