



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 8/10/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 08/09/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 08/10/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115763

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3328-01	20210457-43	1	1.13	3.24	2.02	42.2	sediment sample
M3328-02	20210473-59	2	1.14	5.04	2.9	45.1	sediment sample
M3328-03	20210474-60	3	1.13	9.59	2.45	15.6	sediment sample
M3328-04	20210566-COM-39	4	1.15	9.95	5.49	49.3	sediment sample
M3331-01	BSY1-SB-10-1-2	5	1.14	9.93	9.08	90.3	
M3331-02	BSY1-SB-10-3-4	6	1.16	9.83	8.71	87.1	
M3331-03	BSY1-SB-11-0-1	7	1.17	9.62	8.85	90.9	
M3331-04	BSY1-SB-11-9-10	8	1.18	9.64	9.22	95.0	
M3331-05	BSY1-SB-12-0-1	9	1.13	9.65	8.93	91.5	
M3331-06	BSY1-SB-12-2-3	10	1.11	9.76	8.61	86.7	
M3331-07	BSY1-SB-1011-0-10	11	1.15	9.97	9.16	90.8	
M3331-08	BSY1-SB-13-0-1	12	1.13	9.75	7.95	79.1	
M3331-09	BSY1-SB-13-1-2	13	1.16	9.85	8.49	84.3	
M3331-10	BSY1-SB-14-0-1	14	1.14	9.95	8.36	82.0	
M3331-11	BSY1-SB-14-2-3	15	1.11	9.71	8.58	86.9	
M3331-12	BSY1-SB-121314-0-3	16	1.15	9.98	9.19	91.1	
M3331-13	BSY1-WC-S-10-14-0-10	17	1.12	9.68	8.73	88.9	
M3333-01	124055	18	1.13	9.76	6.47	61.9	
M3333-03	124030	19	1.12	9.54	6.63	65.4	
M3333-05	124019	20	1.18	9.97	7.55	72.5	
M3333-07	124027	21	1.18	9.96	9.3	92.5	
M3333-09	124035	22	1.16	9.62	7.56	75.7	
M3333-11	124042	23	1.15	9.95	9.37	93.4	
M3334-01	VNJ-235	24	1.18	9.74	8.83	89.4	
M3335-01	OR-02-080921	25	1.17	9.58	8.88	91.7	
M3335-02	OR-02-080921	26	1.14	9.88	9.12	91.3	
M3336-01	HS-STONE	27	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
M3337-01	72-12000	28	1.15	9.59	8.82	90.9	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 8/10/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 08/09/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 08/10/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115763

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3338-01	VNJ-202	29	1.15	9.72	8.62	87.2	
M3338-03	VNJ-223	30	1.09	9.65	9.38	96.8	
M3339-01	OK-01-080921	31	1.11	9.76	9.33	95.0	
M3339-02	OK-01-080921	32	1.18	9.69	9.32	95.7	
M3340-01	205-239-929	33	1.18	9.54	9.21	96.1	
M3340-02	205-239-929	34	1.17	9.75	9.42	96.2	
M3341-01	228-344	35	1.16	9.74	9.31	95.0	
M3341-02	228-344	36	1.15	9.61	9.26	95.9	
M3342-01	72-12017	37	1.15	9.5	9.22	96.6	
M3342-01D	72-12017DUP	38	1.15	9.52	9.25	96.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$