


CB115767

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.8906e-004

C= -3.0440e-002

Rho= 0.9999378

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.024	-0.024	23	0.000	23					-15
0.2	0.200	0.170	-0.030	694	0.0 %	694					1
2.5	2.500	2.519	0.019	8819	0.0 %	8819					1
5.0	5.000	5.069	0.069	17643	0.0 %	17643					0
7.5	7.500	7.521	0.021	26125	0.0 %	26125					-1
10.0	10.000	9.944	-0.056	34508	0.0 %	34508					

LB115767 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 10 Aug 2021 08:30:44					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	7470A	PPB	10 Aug 2021 08:33:40	S	Std
0.2 - 1	50.2	694	-	0.2000	7470A	PPB	10 Aug 2021 08:35:57	S	Std
2.5 - 1	52.5	8819	-	2.5000	7470A	PPB	10 Aug 2021 08:38:14	S	Std
5.0 - 1	55	17643	-	5.0000	7470A	PPB	10 Aug 2021 08:40:31	S	Std
7.5 - 1	57.5	26125	-	7.5000	7470A	PPB	10 Aug 2021 08:42:47	S	Std
10.0 - 1	50	34508	-	10.0000	7470A	PPB	10 Aug 2021 08:45:03	S	Std
ICV45 - 1	ICV45	14197	4.0733	-	7470A	PPB	10 Aug 2021 08:47:44	U	SMPL
ICB45 - 1	ICB45	-86	-0.0553	-	7470A	PPB	10 Aug 2021 08:49:59	U	SMPL
CCV21 - 1	CCV21	16656	4.7841	-	7470A	PPB	10 Aug 2021 08:52:13	U	SMPL
CCB21 - 1	CCB21	-73	-0.0515	-	7470A	PPB	10 Aug 2021 08:54:28	U	SMPL
CRA - 1	CRA	819	0.2063	-	7470A	PPB	10 Aug 2021 09:02:50	U	SMPL
HighStd - 1	HighStd	35176	10.1374	-	7470A	PPB	10 Aug 2021 09:07:37	U	SMPL
ChkStd - 1	ChkStd	24140	6.9474	-	7470A	PPB	10 Aug 2021 09:12:30	U	SMPL
PB138292BL - 1	PBW	-109	-0.0619	-	7470A	PPB	10 Aug 2021 09:14:45	U	SMPL
PB138292BS - 1	LCSW	14136	4.0557	-	7470A	PPB	10 Aug 2021 09:17:01	U	SMPL
M3307-03 - 1	VNJ219-367	-43	-0.0429	-	7470A	PPB	10 Aug 2021 09:19:19	U	SMPL
M3308-02 - 1	ETGI-325	7	-0.0284	-	7470A	PPB	10 Aug 2021 09:21:35	U	SMPL
M3308-05 - 1	72-11930	24	-0.0235	-	7470A	PPB	10 Aug 2021 09:23:52	U	SMPL
M3316-01 - 1	SOIL	18	-0.0252	-	7470A	PPB	10 Aug 2021 09:26:09	U	SMPL
M3316-01DUP - 1	SOILDUP	18	-0.0252	-	7470A	PPB	10 Aug 2021 09:28:25	U	SMPL
CCV22 - 1	CCV22	17002	4.8841	-	7470A	PPB	10 Aug 2021 09:33:26	U	SMPL
CCB22 - 1	CCB22	-45	-0.0434	-	7470A	PPB	10 Aug 2021 09:35:40	U	SMPL
M3316-01MS - 1	SOILMS	13594	3.8990	-	7470A	PPB	10 Aug 2021 09:37:56	U	SMPL
M3316-01MSD - 1	SOILMSD	14200	4.0742	-	7470A	PPB	10 Aug 2021 09:40:11	U	SMPL
M3326-02 - 1	MDE-NB	-55	-0.0463	-	7470A	PPB	10 Aug 2021 09:42:26	U	SMPL
PB138293BL - 1	PBW	15	-0.0261	-	7470A	PPB	10 Aug 2021 09:44:41	U	SMPL
PB138293BS - 1	LCSW	14914	4.2806	-	7470A	PPB	10 Aug 2021 09:46:58	U	SMPL
M3279-10 - 1	SPLP-AMENDED-COM-3	82	-0.0067	-	7470A	PPB	10 Aug 2021 09:49:14	U	SMPL
M3279-10DUP - 1	SPLP-AMENDED-COM-3DUP	124	0.0054	-	7470A	PPB	10 Aug 2021 09:51:30	U	SMPL
M3279-10MS - 1	SPLP-AMENDED-COM-3MS	15117	4.3392	-	7470A	PPB	10 Aug 2021 09:53:47	U	SMPL
M3279-10MSD - 1	SPLP-AMENDED-COM-3MSD	15645	4.4919	-	7470A	PPB	10 Aug 2021 09:56:03	U	SMPL
M3318-05 - 1	SPLP-AMENDED-COM-9	64	-0.0119	-	7470A	PPB	10 Aug 2021 09:58:19	U	SMPL
CCV23 - 1	CCV23	18738	5.3859	-	7470A	PPB	10 Aug 2021 10:00:35	U	SMPL
CCB23 - 1	CCB23	-55	-0.0463	-	7470A	PPB	10 Aug 2021 10:02:51	U	SMPL
PB138294BL - 1	PBW	7	-0.0284	-	7470A	PPB	10 Aug 2021 10:05:06	U	SMPL
PB138294BS - 1	LCSW	14862	4.2655	-	7470A	PPB	10 Aug 2021 10:10:25	U	SMPL
M3261-10 - 1	31102	-51	-0.0452	-	7470A	PPB	10 Aug 2021 10:12:39	U	SMPL
M3261-10DUP - 1	31102DUP	32	-0.0212	-	7470A	PPB	10 Aug 2021 10:14:55	U	SMPL
M3261-10MS - 1	31102MS	14788	4.2441	-	7470A	PPB	10 Aug 2021 10:17:10	U	SMPL
M3261-10MSD - 1	31102MSD	14684	4.2141	-	7470A	PPB	10 Aug 2021 10:19:26	U	SMPL
PB138292TB - 1	PB138250TB	-37	-0.0411	-	7470A	PPB	10 Aug 2021 10:21:43	U	SMPL
PB138293TB - 1	PB138254TB	-35	-0.0406	-	7470A	PPB	10 Aug 2021 10:24:00	U	SMPL
PB138294TB - 1	PB138263TB	15	-0.0261	-	7470A	PPB	10 Aug 2021 10:30:44	U	SMPL
CCV24 - 1	CCV24	17555	5.0439	-	7470A	PPB	10 Aug 2021 10:32:59	U	SMPL
CCB24 - 1	CCB24	-45	-0.0434	-	7470A	PPB	10 Aug 2021 10:35:14	U	SMPL
M3316-01LX5 - 1		-47	-0.0440	-	7470A	PPB	10 Aug 2021 10:37:29	U	SMPL
M3316-01A - 1		16320	4.6870	-	7470A	PPB	10 Aug 2021 10:39:45	U	SMPL
M3279-10LX5 - 1		-59	-0.0475	-	7470A	PPB	10 Aug 2021 10:42:00	U	SMPL
M3279-10A - 1		15241	4.3751	-	7470A	PPB	10 Aug 2021 10:44:15	U	SMPL
M3261-10LX5 - 1		-97	-0.0585	-	7470A	PPB	10 Aug 2021 10:46:31	U	SMPL
M3261-10A - 1		14643	4.2022	-	7470A	PPB	10 Aug 2021 10:48:46	U	SMPL
CCV25 - 1	CCV25	17632	5.0662	-	7470A	PPB	10 Aug 2021 10:51:03	U	SMPL
CCB25 - 1	CCB25	-34	-0.0403	-	7470A	PPB	10 Aug 2021 10:53:19	U	SMPL