



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115768

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/09/2021	17:35
2	CAL2	1	Water	NA	NA	21.3	7.01	08/09/2021	17:36
3	CAL3	1	Water	NA	NA	21.3	10.03	08/09/2021	17:37
4	ICV	1	Water	NA	NA	21.4	7.00	08/09/2021	17:38
5	CCV1	1	Water	NA	NA	21.5	2.01	08/09/2021	17:40
6	M3331-13	1	Solid	20.03	20	23.1	7.51	08/09/2021	17:42
7	M3334-01	1	Solid	20.04	20	24.8	8.10	08/09/2021	17:44
8	M3336-02	1	Solid	20.03	20	23.6	5.45	08/09/2021	17:46
9	M3337-01	1	Solid	20.02	20	23.5	6.52	08/09/2021	17:49
10	M3338-01	1	Solid	20.03	20	23.8	6.70	08/09/2021	17:53
11	M3338-03	1	Solid	20.02	20	22.9	7.16	08/09/2021	18:00
12	M3338-03DUP	1	Solid	20.03	20	22.9	7.17	08/09/2021	18:01
13	CCV2	1	Water	NA	NA	21.6	12.02	08/09/2021	18:05

115768

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3333 WorkList ID : 151278 Department : Wet-Chemistry Date : 08-09-2021 14:59:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/06/2021	Solid	M3331-13	Corrosivity	Cool 4 deg C	PARS15	L11	BSY1-WC-S-10-14-0-10	08/06/2021	9045D
08/09/2021	Solid	M3334-01	Corrosivity	Cool 4 deg C	PSEG03	R42	VNJ-235	08/09/2021	9045D
08/09/2021	Solid	M3336-02	Corrosivity	Cool 4 deg C	PSEG03	R32	HS-STONE	08/09/2021	9045D
08/09/2021	Solid	M3337-01	Corrosivity	Cool 4 deg C	PSEG03	R41	72-12000	08/09/2021	9045D
08/09/2021	Solid	M3338-01	Corrosivity	Cool 4 deg C	PSEG03	R41	VNJ-202	08/09/2021	9045D
08/09/2021	Solid	M3338-03	Corrosivity	Cool 4 deg C	PSEG03	R41	VNJ-223	08/09/2021	9045D

Date/Time 08/09/21 16:10
 Raw Sample Received by: JP (YDC)
 Raw Sample Relinquished by: JP Sin

Date/Time 08/09/21 18:20
 Raw Sample Received by: JP Sin
 Raw Sample Relinquished by: JP (WCI)