


LB115771

7471B

INSTRUMENT ID: CV2

 μ Abs.:

52778



Linear

A= 0.0000e+000

B= 1.8883e-004

C= 2.6280e-002

Rho= 0.9999489

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.035	0.035	45	0.000	45					
0.2	0.200	0.196	-0.004	899	0.0 %	899					-2
2.5	2.500	2.482	-0.018	13004	0.0 %	13004					-1
5.0	5.000	4.939	-0.061	26018	0.0 %	26018					-1
7.5	7.500	7.555	0.055	39871	0.0 %	39871					1
10.0	10.000	9.993	-0.007	52778	0.0 %	52778					0

LB115771 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Aug 2021 11:26:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.0000	7471B	PPB	10 Aug 2021 11:27:30	S	Std
0.2 - 1	50.2	899	-	0.2000	7471B	PPB	10 Aug 2021 11:29:47	S	Std
2.5 - 1	52.5	13004	-	2.5000	7471B	PPB	10 Aug 2021 11:32:04	S	Std
5.0 - 1	55	26018	-	5.0000	7471B	PPB	10 Aug 2021 11:34:20	S	Std
7.5 - 1	57.5	39871	-	7.5000	7471B	PPB	10 Aug 2021 11:36:36	S	Std
10.0 - 1	510	52778	-	10.0000	7471B	PPB	10 Aug 2021 11:41:31	S	Std
ICV47 - 1	ICV47	20257	3.8515	-7471B	PPB	PPB	10 Aug 2021 11:44:09	U	SMPL
ICB47 - 1	ICB47	-49	0.0170	-7471B	PPB	PPB	10 Aug 2021 11:46:23	U	SMPL
CCV25 - 1	CCV25	26135	4.9615	-7471B	PPB	PPB	10 Aug 2021 11:48:38	U	SMPL
CCB25 - 1	CCB25	-54	0.0161	-7471B	PPB	PPB	10 Aug 2021 11:50:53	U	SMPL
CRA - 1	CRA	888	0.1940	-7471B	PPB	PPB	10 Aug 2021 11:53:08	U	SMPL
HighStd - 1	HighStd	52317	9.9055	-7471B	PPB	PPB	10 Aug 2021 11:55:23	U	SMPL
ChkStd - 1	ChkStd	37587	7.1240	-7471B	PPB	PPB	10 Aug 2021 11:57:39	U	SMPL
PB138303BL - 1	PB138303BL	-57	0.0155	-7471B	PPB	PPB	10 Aug 2021 11:59:55	U	SMPL
PB138303BS - 1	PB138303BS	21515	4.0891	-7471B	PPB	PPB	10 Aug 2021 12:02:13	U	SMPL
M3299-01 - 1	20210560-COM-33	22673	4.3077	-7471B	PPB	PPB	10 Aug 2021 12:04:29	U	SMPL
M3299-02 - 1	20210562-COM-35	49980	9.4642	-7471B	PPB	PPB	10 Aug 2021 12:06:46	U	SMPL
M3300-01 - 1	21188	115	0.0480	-7471B	PPB	PPB	10 Aug 2021 12:09:03	U	SMPL
M3305-01 - 1	SA-01-080521	1517	0.3127	-7471B	PPB	PPB	10 Aug 2021 12:11:21	U	SMPL
M3307-01 - 1	VNJ219-367	4375	0.8524	-7471B	PPB	PPB	10 Aug 2021 12:13:37	U	SMPL
CCV26 - 1	CCV26	26128	4.9602	-7471B	PPB	PPB	10 Aug 2021 12:15:52	U	SMPL
CCB26 - 1	CCB26	-43	0.0182	-7471B	PPB	PPB	10 Aug 2021 12:18:08	U	SMPL
M3308-01 - 1	ETGI-325	443	0.1099	-7471B	PPB	PPB	10 Aug 2021 12:20:22	U	SMPL
M3308-04 - 1	72-11930	1783	0.3630	-7471B	PPB	PPB	10 Aug 2021 12:22:38	U	SMPL
M3310-01 - 1	NB-08-080521	2440	0.4870	-7471B	PPB	PPB	10 Aug 2021 12:24:53	U	SMPL
M3314-01 - 1	72-11931	5381	1.0424	-7471B	PPB	PPB	10 Aug 2021 12:27:08	U	SMPL
M3315-01 - 1	OK-02-08062021	3406	0.6695	-7471B	PPB	PPB	10 Aug 2021 12:29:24	U	SMPL
M3318-01 - 1	20210536-COM-9	66525	12.5885	-7471B	PPB	PPB	10 Aug 2021 12:31:40	U	SMPL
M3318-02 - 1	20210540-COM-13	65749	12.4420	-7471B	PPB	PPB	10 Aug 2021 12:33:57	U	SMPL
M3318-03 - 1	20210586-COM-9	54519	10.3214	-7471B	PPB	PPB	10 Aug 2021 12:36:16	U	SMPL
M3318-04 - 1	20210595-AMENDED-COM-9	41036	7.7753	-7471B	PPB	PPB	10 Aug 2021 12:38:35	U	SMPL
CCV27 - 1	CCV27	26019	4.9396	-7471B	PPB	PPB	10 Aug 2021 12:40:52	U	SMPL
CCB27 - 1	CCB27	-56	0.0157	-7471B	PPB	PPB	10 Aug 2021 12:43:08	U	SMPL
M3321-01 - 1	CARLSTADT	3466	0.6808	-7471B	PPB	PPB	10 Aug 2021 12:45:24	U	SMPL
M3321-01DUP - 1	CARLSTADTDUP	3789	0.7418	-7471B	PPB	PPB	10 Aug 2021 12:47:39	U	SMPL
M3321-01MS - 1	CARLSTADTMS	24470	4.6471	-7471B	PPB	PPB	10 Aug 2021 12:49:55	U	SMPL
M3321-01MSD - 1	CARLSTADTMSD	23947	4.5483	-7471B	PPB	PPB	10 Aug 2021 12:52:10	U	SMPL
M3326-01 - 1	MDE-NB	7874	1.5132	-7471B	PPB	PPB	10 Aug 2021 12:54:26	U	SMPL
M3321-01LX5 - 1		669	0.1526	-7471B	PPB	PPB	10 Aug 2021 12:56:41	U	SMPL
M3321-01A - 1		24159	4.5883	-7471B	PPB	PPB	10 Aug 2021 12:58:58	U	SMPL
M3318-01DLX2 - 1		33776	6.4044	-7471B	PPB	PPB	10 Aug 2021 13:01:15	U	SMPL
M3318-02DLX2 - 1		32737	6.2082	-7471B	PPB	PPB	10 Aug 2021 13:03:31	U	SMPL
M3318-03DLX2 - 1		27340	5.1890	-7471B	PPB	PPB	10 Aug 2021 13:05:49	U	SMPL
CCV28 - 1	CCV28	26487	5.0279	-7471B	PPB	PPB	10 Aug 2021 13:08:05	U	SMPL
CCB28 - 1	CCB28	-39	0.0189	-7471B	PPB	PPB	10 Aug 2021 13:10:21	U	SMPL