



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115776

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/10/2021	15:05
2	CAL2	1	Water	NA	NA	21.3	7.01	08/10/2021	15:06
3	CAL3	1	Water	NA	NA	21.3	10.03	08/10/2021	15:08
4	ICV	1	Water	NA	NA	21.4	6.98	08/10/2021	15:10
5	CCV1	1	Water	NA	NA	21.4	2.01	08/10/2021	15:11
6	M3348-01	1	Solid	20.03	20	24.2	5.46	08/10/2021	15:15
7	M3348-01DUP	1	Solid	20.04	20	24.3	5.47	08/10/2021	15:16
8	CCV2	1	Water	NA	NA	21.5	12.03	08/10/2021	15:19

115776

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3348 WorkList ID : 151303 Department : Wet-Chemistry Date : 08-10-2021 14:58:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/10/2021	Solid	M3348-01	pH	Cool 4 deg C	PSEG03	D42	SPEEDIDRI-TRE-1237	08/10/2021	9045D

Date/Time 08-10-21 15:00
Raw Sample Received by: JR (WC)
Raw Sample Relinquished by: CR SM

Date/Time 08-10-21 16:00
Raw Sample Received by: CR SM
Raw Sample Relinquished by: JR (WC)