



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115777

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/10/2021	16:15
2	CAL2	1	Water	NA	NA	21.2	7.00	08/10/2021	16:16
3	CAL3	1	Water	NA	NA	21.3	10.03	08/10/2021	16:17
4	ICV	1	Water	NA	NA	21.4	7.02	08/10/2021	16:19
5	CCV1	1	Water	NA	NA	21.4	2.01	08/10/2021	16:20
6	M3351-01	1	Water	NA	NA	22.2	6.47	08/10/2021	16:21
7	M3351-01DUP	1	Water	NA	NA	22.3	6.48	08/10/2021	16:22
8	CCV2	1	Water	NA	NA	21.6	12.02	08/10/2021	16:25

VB 115777

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3351 WorkList ID : 151308 Department : Wet-Chemistry Date : 08-10-2021 16:08:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/09/2021	Water	M3351-01	pH		SCIE01	D11	SUPPORT-WASH-WAST	08/09/2021	SM 4500-H B

Date/Time 08/10/21 16:10
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)

Date/Time 08-10-21 17:15
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)