


LB115783

245.1

INSTRUMENT 115783



A= 0.0000e+000

B= 2.7493e-004

C= 1.9643e-002

Rho= 0.9999389

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/1
0.0	0.000	0.009	0.009	-38	0.000	-38					4
0.05	0.050	0.062	0.012	153	0.0 %	153	0				-17
0.2	0.200	0.166	-0.034	531	0.0 %	531					1
2.5	2.500	2.464	-0.036	8892	0.0 %	8892					1
5.0	5.000	5.056	0.056	18318	0.0 %	18318					1
7.5	7.500	7.551	0.051	27394	0.0 %	27394					1
10.0	10.000	9.942	-0.058	36090	0.0 %	36090					1

LB115783 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Aug 2021 11:31:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SB	-38	-	0.0000	245.1	PPB	11 Aug 2021 11:33:42	S	Std
0.05 - 1	50.05	153	-	0.0500	245.1	PPB	11 Aug 2021 11:35:59	S	Std
0.2 - 1	10.2	531	-	0.2000	245.1	PPB	11 Aug 2021 11:38:16	S	Std
2.5 - 1	52.5	8892	-	2.5000	245.1	PPB	11 Aug 2021 11:40:32	S	Std
5.0 - 1	15	18318	-	5.0000	245.1	PPB	11 Aug 2021 11:42:49	S	Std
7.5 - 1	57.5	27394	-	7.5000	245.1	PPB	11 Aug 2021 11:45:04	S	Std
10.0 - 1	51.0	36090	-	10.0000	245.1	PPB	11 Aug 2021 11:54:54	S	Std
ICV48 - 1	ICV48	13885	3.8371	-245.1		PPB	11 Aug 2021 12:00:47	U	SMPL
ICB48 - 1	ICB48	-111	-0.0109	-245.1		PPB	11 Aug 2021 12:03:02	U	SMPL
CCV29 - 1	CCV29	18346	5.0636	-245.1		PPB	11 Aug 2021 12:05:17	U	SMPL
CCB29 - 1	CCB29	-123	-0.0142	-245.1		PPB	11 Aug 2021 12:07:32	U	SMPL
CRA - 1	CRA	153	0.0617	-245.1		PPB	11 Aug 2021 12:09:47	U	SMPL
HighStd - 1	HighStd	36133	9.9538	-245.1		PPB	11 Aug 2021 12:12:01	U	SMPL
ChkStd - 1	ChkStd	25510	7.0332	-245.1		PPB	11 Aug 2021 12:14:17	U	SMPL
PB138331BL - 1	PBW	-144	-0.0199	-245.1		PPB	11 Aug 2021 12:16:34	U	SMPL
PB138331BS - 1	LCSW	14087	3.8926	-245.1		PPB	11 Aug 2021 12:18:49	U	SMPL
M3281-03 - 1	TWP01	1483	0.4274	-245.1		PPB	11 Aug 2021 12:21:05	U	SMPL
M3281-03DUP - 1	TWP01DUP	1380	0.3991	-245.1		PPB	11 Aug 2021 12:23:21	U	SMPL
M3281-03MS - 1	TWP01MS	17407	4.8054	-245.1		PPB	11 Aug 2021 12:25:38	U	SMPL
M3281-03MSD - 1	TWP01MSD	17014	4.6974	-245.1		PPB	11 Aug 2021 12:27:56	U	SMPL
M3281-04 - 1	TWP01	78	0.0411	-245.1		PPB	11 Aug 2021 12:30:11	U	SMPL
CCV30 - 1	CCV30	18299	5.0507	-245.1		PPB	11 Aug 2021 12:32:27	U	SMPL
CCB30 - 1	CCB30	-110	-0.0106	-245.1		PPB	11 Aug 2021 12:34:42	U	SMPL
M3281-05 - 1	TWP02	345	0.1145	-245.1		PPB	11 Aug 2021 12:36:57	U	SMPL
M3281-06 - 1	TWP02	88	0.0438	-245.1		PPB	11 Aug 2021 12:39:12	U	SMPL
M3281-07 - 1	TWP03	980	0.2891	-245.1		PPB	11 Aug 2021 12:41:28	U	SMPL
M3281-08 - 1	TWP03	18	0.0246	-245.1		PPB	11 Aug 2021 12:43:44	U	SMPL
M3285-01 - 1	GRAB	375	0.1227	-245.1		PPB	11 Aug 2021 12:46:00	U	SMPL
M3303-01 - 1	315220012-1	-16	0.0152	-245.1		PPB	11 Aug 2021 12:48:16	U	SMPL
M3281-03LX5 - 1		176	0.0680	-245.1		PPB	11 Aug 2021 12:50:33	U	SMPL
M3281-03A - 1		14744	4.0733	-245.1		PPB	11 Aug 2021 12:52:50	U	SMPL
CCV31 - 1	CCV31	18171	5.0155	-245.1		PPB	11 Aug 2021 12:55:07	U	SMPL
CCB31 - 1	CCB31	-114	-0.0117	-245.1		PPB	11 Aug 2021 12:57:22	U	SMPL