


LB115984

7471B

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 1.9548e-004

C= 4.5030e-003

Rho= 0.9999740

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.035	0.035	156	0.000	156					1
0.2	0.200	0.201	0.001	1003	0.0 %	1003					1
2.5	2.500	2.483	-0.017	12681	0.0 %	12681					1
5.0	5.000	4.961	-0.039	25355	0.0 %	25355					0
7.5	7.500	7.487	-0.013	38279	0.0 %	38279					0
10.0	10.000	10.033	0.033	51300	0.0 %	51300					0

LB115784 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 11 Aug 2021 13:04:58				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	SD	156	-	0.0000	7471B	PPB	11 Aug 2021 13:06:04	S Std
0.2 - 1	SD	1003	-	0.2000	7471B	PPB	11 Aug 2021 13:08:21	S Std
2.5 - 1	SD	12681	-	2.5000	7471B	PPB	11 Aug 2021 13:10:38	S Std
5.0 - 1	SD	25355	-	5.0000	7471B	PPB	11 Aug 2021 13:12:54	S Std
7.5 - 1	SD	38279	-	7.5000	7471B	PPB	11 Aug 2021 13:15:11	S Std
10.0 - 1	SD	51300	-	10.0000	7471B	PPB	11 Aug 2021 13:17:27	S Std
ICV49 - 1	ICV49	20050	3.9239	-	7471B	PPB	11 Aug 2021 13:20:15	U SMPL
ICB49 - 1	ICB49	-49	-0.0051	-	7471B	PPB	11 Aug 2021 13:22:29	U SMPL
CCV32 - 1	CCV32	26194	5.1250	-	7471B	PPB	11 Aug 2021 13:24:44	U SMPL
CCB32 - 1	CCB32	-59	-0.0070	-	7471B	PPB	11 Aug 2021 13:26:59	U SMPL
CRA - 1	CRA	987	0.1974	-	7471B	PPB	11 Aug 2021 13:29:13	U SMPL
HighStd - 1	HighStd	51107	9.9950	-	7471B	PPB	11 Aug 2021 13:31:28	U SMPL
ChkStd - 1	ChkStd	36202	7.0813	-	7471B	PPB	11 Aug 2021 13:33:43	U SMPL
PB138333BL - 1	PBS	-31	-0.0016	-	7471B	PPB	11 Aug 2021 13:36:01	U SMPL
PB138333BS - 1	LCSS	21349	4.1778	-	7471B	PPB	11 Aug 2021 13:38:18	U SMPL
M3328-04 - 1	20210566-COM-39	31099	6.0838	-	7471B	PPB	11 Aug 2021 13:40:34	U SMPL
M3331-07 - 1	BSY1-SB-1011-0-10	3332	0.6558	-	7471B	PPB	11 Aug 2021 13:47:48	U SMPL
M3331-07DUP - 1	BSY1-SB-1011-0-10DUP	3155	0.6212	-	7471B	PPB	11 Aug 2021 13:50:05	U SMPL
M3331-07MS - 1	BSY1-SB-1011-0-10MS	24503	4.7944	-	7471B	PPB	11 Aug 2021 13:52:22	U SMPL
M3331-07MSD - 1	BSY1-SB-1011-0-10MSD	22156	4.3356	-	7471B	PPB	11 Aug 2021 13:54:37	U SMPL
CCV33 - 1	CCV33	26155	5.1173	-	7471B	PPB	11 Aug 2021 13:56:52	U SMPL
CCB33 - 1	CCB33	-39	-0.0031	-	7471B	PPB	11 Aug 2021 13:59:08	U SMPL
M3331-12 - 1	BSY1-SB-121314-0-3	7173	1.4067	-	7471B	PPB	11 Aug 2021 14:01:23	U SMPL
M3343-01 - 1	20210542-COM-15	69000	13.4928	-	7471B	PPB	11 Aug 2021 14:03:39	U SMPL
M3343-02 - 1	20210544-COM-17	48783	9.5407	-	7471B	PPB	11 Aug 2021 14:05:54	U SMPL
M3343-03 - 1	20210547-COM-20	83629	16.3525	-	7471B	PPB	11 Aug 2021 14:08:11	U SMPL
M3343-04 - 1	20210548-COM-21	37986	7.4301	-	7471B	PPB	11 Aug 2021 14:10:28	U SMPL
M3343-05 - 1	20210553-COM-26	31064	6.0770	-	7471B	PPB	11 Aug 2021 14:12:46	U SMPL
M3343-06 - 1	20210555-COM-28	36218	7.0845	-	7471B	PPB	11 Aug 2021 14:15:03	U SMPL
M3343-07 - 1	20210558-COM-31	49683	9.7166	-	7471B	PPB	11 Aug 2021 14:17:20	U SMPL
M3343-08 - 1	20210582-COM-5	48698	9.5241	-	7471B	PPB	11 Aug 2021 14:19:39	U SMPL
M3343-09 - 1	20210591-AMENDED-COM-5	30682	6.0023	-	7471B	PPB	11 Aug 2021 14:21:56	U SMPL
CCV34 - 1	CCV34	26260	5.1379	-	7471B	PPB	11 Aug 2021 14:24:13	U SMPL
CCB34 - 1	CCB34	-46	-0.0045	-	7471B	PPB	11 Aug 2021 14:26:30	U SMPL
PB138334BL - 1	PBS	40	0.0123	-	7471B	PPB	11 Aug 2021 14:28:45	U SMPL
PB138334BS - 1	LCSS	21291	4.1665	-	7471B	PPB	11 Aug 2021 14:31:01	U SMPL
M3333-01 - 1	124055	2784	0.5487	-	7471B	PPB	11 Aug 2021 14:33:17	U SMPL
M3333-03 - 1	124030	2337	0.4613	-	7471B	PPB	11 Aug 2021 14:35:32	U SMPL
M3333-05 - 1	124019	4577	0.8992	-	7471B	PPB	11 Aug 2021 14:37:48	U SMPL
M3333-07 - 1	124027	3570	0.7024	-	7471B	PPB	11 Aug 2021 14:40:04	U SMPL
M3333-09 - 1	124035	3306	0.6508	-	7471B	PPB	11 Aug 2021 14:42:21	U SMPL
M3333-11 - 1	124042	1300	0.2586	-	7471B	PPB	11 Aug 2021 14:44:38	U SMPL
M3334-01 - 1	VNJ-235	3459	0.6807	-	7471B	PPB	11 Aug 2021 14:46:54	U SMPL
M3335-01 - 1	OR-02-080921	1539	0.3053	-	7471B	PPB	11 Aug 2021 14:49:10	U SMPL
CCV35 - 1	CCV35	26465	5.1779	-	7471B	PPB	11 Aug 2021 14:51:25	U SMPL
CCB35 - 1	CCB35	-106	-0.0162	-	7471B	PPB	11 Aug 2021 14:53:41	U SMPL
M3336-01 - 1	HS-STONE	160	0.0358	-	7471B	PPB	11 Aug 2021 14:55:56	U SMPL
M3339-01 - 1	OK-01-080921	1504	0.2985	-	7471B	PPB	11 Aug 2021 14:58:12	U SMPL
M3340-01 - 1	205-239-929	2098	0.4146	-	7471B	PPB	11 Aug 2021 15:00:27	U SMPL
M3341-01 - 1	228-344	1031	0.2060	-	7471B	PPB	11 Aug 2021 15:02:42	U SMPL
M3341-01DUP - 1	228-344DUP	964	0.1929	-	7471B	PPB	11 Aug 2021 15:04:58	U SMPL
M3341-01MS - 1	228-344MS	21585	4.2240	-	7471B	PPB	11 Aug 2021 15:07:14	U SMPL
M3341-01MSD - 1	228-344MSD	21202	4.1491	-	7471B	PPB	11 Aug 2021 15:12:08	U SMPL
M3346-01 - 1	LPA-1	3495	0.6877	-	7471B	PPB	11 Aug 2021 15:14:24	U SMPL
CCV36 - 1	CCV36	26140	5.1144	-	7471B	PPB	11 Aug 2021 15:23:22	U SMPL
CCB36 - 1	CCB36	-40	-0.0033	-	7471B	PPB	11 Aug 2021 15:25:38	U SMPL
M3331-07LX5 - 1		174	0.0385	-	7471B	PPB	11 Aug 2021 15:27:53	U SMPL
M3331-07A - 1		23281	4.5555	-	7471B	PPB	11 Aug 2021 15:30:08	U SMPL
M3341-01LX5 - 1		66	0.0174	-	7471B	PPB	11 Aug 2021 15:32:24	U SMPL
M3341-01A - 1		21645	4.2357	-	7471B	PPB	11 Aug 2021 15:34:39	U SMPL
M3343-01DLX2 - 1		34341	6.7175	-	7471B	PPB	11 Aug 2021 15:36:54	U SMPL
M3343-03DLX5 - 1		17316	3.3895	-	7471B	PPB	11 Aug 2021 15:39:10	U SMPL
CCV37 - 1	CCV37	25659	5.0204	-	7471B	PPB	11 Aug 2021 15:41:27	U SMPL
CCB37 - 1	CCB37	9	0.0063	-	7471B	PPB	11 Aug 2021 15:43:43	U SMPL