

Analysis Method: SM2510 BANALYST: apatelParameter: ConductanceSUPERVISOR REVIEW BY: ketankumarRun Number: LB115786Meter ID: WC Conductance Meter-1Thermometer ID: WC pH meter probe-1

Reagent/Standard	Lot/Log #
conductivity standard	WP89507

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB115786BL		1	24.2	0.82	0.83	08/11/2021	12:52
2	LB115786BS	1413	1	24.3	1398.00	1416.94	08/11/2021	12:55
3	M3367-01		1	24.1	1.35	1.37	08/11/2021	13:00
4	M3367-01DUP		1	24.2	1.37	1.39	08/11/2021	13:05

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Cond-081121 WorkList ID : 151329 Department : Wet-Chemistry Date : 08-11-2021 11:55:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/24/2021	Water	M3367-01	Conductance	Cool 4 deg C	STER06	D42	POOL-WATER	08/10/2021	SM2510 B

Date/Time 8/11/21 12:45
Raw Sample Received by: AB(WC)
Raw Sample Relinquished by: JP(WC)

Date/Time 8/11/21 13:10
Raw Sample Received by: JP(WC)
Raw Sample Relinquished by: AB(WC)