



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115787

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/11/2021	13:45
2	CAL2	1	Water	NA	NA	21.3	7.00	08/11/2021	13:46
3	CAL3	1	Water	NA	NA	21.3	10.03	08/11/2021	13:47
4	ICV	1	Water	NA	NA	21.4	6.99	08/11/2021	13:50
5	CCV1	1	Water	NA	NA	21.3	2.01	08/11/2021	14:00
6	M3366-01	1	Water	NA	NA	23.0	5.77	08/11/2021	14:01
7	M3366-02	1	Water	NA	NA	23.2	5.55	08/11/2021	14:05
8	M3366-03	1	Water	NA	NA	23.3	5.43	08/11/2021	14:09
9	M3367-01	1	Water	NA	NA	23.5	5.89	08/11/2021	14:13
10	M3367-01DUP	1	Water	NA	NA	23.6	5.90	08/11/2021	14:14
11	CCV2	1	Water	NA	NA	21.4	12.03	08/11/2021	14:15

WORKLIST(Hardcopy Internal Chain)

LB15787

WorkList Name : PH M3366 WorkList ID : 151330 Department : Wet-Chemistry Date : 08-11-2021 13:35:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/09/2021	Water	M3366-01	pH	Cool 4 deg C	STER13	D11	CELL-A-POOL-WATER	08/09/2021	9040C
08/09/2021	Water	M3366-02	pH	Cool 4 deg C	STER13	D11	CELL-B-POOL-WATER	08/09/2021	9040C
08/09/2021	Water	M3366-03	pH	Cool 4 deg C	STER13	D11	CELL-C-POOL-WATER	08/09/2021	9040C
08/10/2021	Water	M3367-01	pH	Cool 4 deg C	STER06	D42	POOL-WATER	08/10/2021	9040C

Date/Time 08/11/21 13:40
 Raw Sample Received by: JR (WC)
 Raw Sample Relinquished by: JR (WC)

Date/Time 08/11/21 14:40
 Raw Sample Received by: JR (WC)
 Raw Sample Relinquished by: JR (WC)