



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115794

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	08/11/2021	17:00
2	CAL2	1	Water	NA	NA	22.1	7.01	08/11/2021	17:01
3	CAL3	1	Water	NA	NA	22.1	10.02	08/11/2021	17:02
4	ICV	1	Water	NA	NA	22.1	6.99	08/11/2021	17:05
5	CCV1	1	Water	NA	NA	22.1	2.01	08/11/2021	17:06
6	M3346-02	1	Solid	20.03	20	24.3	8.36	08/11/2021	17:09
7	M3371-02	1	Solid	20.02	20	24.1	10.12	08/11/2021	17:10
8	M3371-02DUP	1	Solid	20.03	20	24.2	10.13	08/11/2021	17:11
9	CCV2	1	Water	NA	NA	22.2	12.02	08/11/2021	17:12

MP 115794

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3371

WorkList ID : 151335

Department : Wet-Chemistry

Date : 08-11-2021 16:39:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/10/2021	Solid	M3346-02	☒ Corrosivity	Cool 4 deg C	PSEG03	D42	LPA-1	08/10/2021	9045D
08/10/2021	Solid	M3371-02	☒ Corrosivity	Cool 4 deg C	TULL02	D11	JET-GROUT	08/10/2021	9045D

Date/Time08-11-2116:40

Raw Sample Received by:JP (WC)

Raw Sample Relinquished by:JP (WC)

Date/Time08-11-2117:50

Raw Sample Received by:JP (WC)

Raw Sample Relinquished by:JP (WC)