



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115798

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.2	4.01	08/12/2021	08:05
2	CAL2	1	Water	NA	NA	22.2	7.00	08/12/2021	08:06
3	CAL3	1	Water	NA	NA	22.2	10.03	08/12/2021	08:07
4	ICV	1	Water	NA	NA	22.3	7.01	08/12/2021	08:10
5	CCV1	1	Water	NA	NA	22.3	2.01	08/12/2021	08:15
6	M3360-01	1	Water	NA	NA	24.5	6.92	08/12/2021	08:19
7	M3360-02	1	Water	NA	NA	25.0	6.83	08/12/2021	08:22
8	M3360-03	1	Water	NA	NA	24.6	7.08	08/12/2021	08:25
9	M3360-04	1	Water	NA	NA	24.6	6.91	08/12/2021	08:29
10	M3360-05	1	Water	NA	NA	25.0	6.63	08/12/2021	08:30
11	M3361-07	1	Water	NA	NA	28.5	6.43	08/12/2021	08:31
12	M3361-08	1	Water	NA	NA	26.1	5.39	08/12/2021	08:32
13	M3361-09	1	Water	NA	NA	26.7	6.88	08/12/2021	08:35
14	M3361-10	1	Water	NA	NA	25.2	7.85	08/12/2021	08:36
15	M3361-11	1	Water	NA	NA	26.2	7.61	08/12/2021	08:38
16	CCV2	1	Water	NA	NA	22.3	12.02	08/12/2021	08:40
17	M3361-12	1	Water	NA	NA	25.2	6.20	08/12/2021	08:42
18	M3361-13	1	Water	NA	NA	25.9	4.11	08/12/2021	08:45
19	M3361-14	1	Water	NA	NA	25.8	3.90	08/12/2021	08:49
20	M3361-18	1	Water	NA	NA	26.4	6.53	08/12/2021	08:52

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M3361-18DUP	1	Water	NA	NA	26.5	6.54	08/12/2021	08:53
22	CCV3	1	Water	NA	NA	22.6	2.01	08/12/2021	09:00

115198

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3360 WorkList ID : 151339 Department : Wet-Chemistry Date : 08-12-2021 07:55:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/11/2021	Water	M3360-01	✓ pH	Cool 4 deg C	PSEG03	D42	41266	08/11/2021	9040C
08/11/2021	Water	M3360-02	✓ pH	Cool 4 deg C	PSEG03	D42	41265	08/11/2021	9040C
08/11/2021	Water	M3360-03	✓ pH	Cool 4 deg C	PSEG03	D42	41920	08/11/2021	9040C
08/11/2021	Water	M3360-04	✓ pH	Cool 4 deg C	PSEG03	D42	41935	08/11/2021	9040C
08/11/2021	Water	M3360-05	✓ pH	Cool 4 deg C	PSEG03	D42	31109	08/11/2021	9040C
08/11/2021	Water	M3361-07	✓ pH	Cool 4 deg C	PSEG03	D42	41907	08/11/2021	9040C
08/11/2021	Water	M3361-08	✓ pH	Cool 4 deg C	PSEG03	D42	41908	08/11/2021	9040C
08/11/2021	Water	M3361-09	✓ pH	Cool 4 deg C	PSEG03	D42	41911	08/11/2021	9040C
08/11/2021	Water	M3361-10	✓ pH	Cool 4 deg C	PSEG03	D42	41912	08/11/2021	9040C
08/11/2021	Water	M3361-11	✓ pH	Cool 4 deg C	PSEG03	D42	41942	08/11/2021	9040C
08/11/2021	Water	M3361-12	✓ pH	Cool 4 deg C	PSEG03	D42	41940	08/11/2021	9040C
08/11/2021	Water	M3361-13	✓ pH	Cool 4 deg C	PSEG03	D42	41943	08/11/2021	9040C
08/11/2021	Water	M3361-14	✓ pH	Cool 4 deg C	PSEG03	D42	41909	08/11/2021	9040C
08/11/2021	Water	M3361-18	✓ pH	Cool 4 deg C	PSEG03	D42	41949	08/11/2021	9040C

Date/Time 08/12/21 08:00

Raw Sample Received by: JP (WC)

Raw Sample Relinquished by: JP (WC)

Date/Time 08/12/21 10:00

Raw Sample Received by: JP (WC)

Raw Sample Relinquished by: JP (WC)