

LB 115799


7470A

INSTRUMENT 10: CV

Linear

 μ Abs.:

41337



A= 0.0000e+000

B= 2.4115e-004

C= 3.5789e-002

Rho= 0.9999823

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.034	0.034	-6	0.000	-6					-
0.2	0.200	0.189	-0.011	636	0.0 %	636					-6
2.5	2.500	2.482	-0.018	10142	0.0 %	10142					-1
5.0	5.000	4.971	-0.029	20465	0.0 %	20465					-1
7.5	7.500	7.520	0.020	31035	0.0 %	31035					0
10.0	10.000	10.004	0.004	41337	0.0 %	41337					0

LB115799 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Aug 2021 09:05:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-6	-	0.0000	7470A	PPB	12 Aug 2021 09:12:20	S	Std
0.2 - 1	50-2	636	-	0.2000	7470A	PPB	12 Aug 2021 09:14:37	S	Std
2.5 - 1	52-5	10142	-	2.5000	7470A	PPB	12 Aug 2021 09:16:54	S	Std
5.0 - 1	55	20465	-	5.0000	7470A	PPB	12 Aug 2021 09:19:10	S	Std
7.5 - 1	57-5	31035	-	7.5000	7470A	PPB	12 Aug 2021 09:21:26	S	Std
10.0 - 1	510	41337	-	10.0000	7470A	PPB	12 Aug 2021 09:26:16	S	Std
ICV50 - 1	ICV50	15138	3.6863	-	7470A	PPB	12 Aug 2021 09:29:01	U	SMPL
ICB50 - 1	ICB50	-35	0.0273	-	7470A	PPB	12 Aug 2021 09:33:47	U	SMPL
CCV38 - 1	CCV38	20644	5.0141	-	7470A	PPB	12 Aug 2021 09:36:02	U	SMPL
CCB38 - 1	CCB38	-106	0.0102	-	7470A	PPB	12 Aug 2021 09:38:17	U	SMPL
CRA - 1	CRA	625	0.1865	-	7470A	PPB	12 Aug 2021 09:40:31	U	SMPL
HighStd - 1	HighStd	41411	10.0220	-	7470A	PPB	12 Aug 2021 09:42:47	U	SMPL
ChkStd - 1	ChkStd	28571	6.9257	-	7470A	PPB	12 Aug 2021 09:45:01	U	SMPL
PB138313BL - 1	PBW	-143	0.0013	-	7470A	PPB	12 Aug 2021 09:47:18	U	SMPL
PB138313BS - 1	LCSW	15845	3.8568	-	7470A	PPB	12 Aug 2021 09:49:34	U	SMPL
M3331-13 - 1	BSY1-WC-S-10-14-0-10	-86	0.0151	-	7470A	PPB	12 Aug 2021 09:51:51	U	SMPL
M3331-13DUP - 1	BSY1-WC-S-10-14-0-10DUP	-11	0.0331	-	7470A	PPB	12 Aug 2021 09:54:07	U	SMPL
M3331-13MS - 1	BSY1-WC-S-10-14-0-10MS	15456	3.7630	-	7470A	PPB	12 Aug 2021 09:56:24	U	SMPL
M3331-13MSD - 1	BSY1-WC-S-10-14-0-10MSD	15941	3.8800	-	7470A	PPB	12 Aug 2021 09:58:41	U	SMPL
M3333-02 - 1	1245055	-67	0.0196	-	7470A	PPB	12 Aug 2021 10:00:57	U	SMPL
CCV39 - 1	CCV39	20723	5.0331	-	7470A	PPB	12 Aug 2021 10:03:13	U	SMPL
CCB39 - 1	CCB39	-110	0.0093	-	7470A	PPB	12 Aug 2021 10:05:28	U	SMPL
M3333-04 - 1	124030	16	0.0396	-	7470A	PPB	12 Aug 2021 10:07:44	U	SMPL
M3333-06 - 1	124019	-4	0.0348	-	7470A	PPB	12 Aug 2021 10:10:00	U	SMPL
M3333-08 - 1	124027	11	0.0384	-	7470A	PPB	12 Aug 2021 10:12:15	U	SMPL
M3333-10 - 1	124035	-10	0.0334	-	7470A	PPB	12 Aug 2021 10:14:30	U	SMPL
M3333-12 - 1	124042	-11	0.0331	-	7470A	PPB	12 Aug 2021 10:16:46	U	SMPL
M3334-02 - 1	VNJ-235	0	0.0358	-	7470A	PPB	12 Aug 2021 10:19:02	U	SMPL
M3336-02 - 1	HS-STONE	3	0.0365	-	7470A	PPB	12 Aug 2021 10:21:18	U	SMPL
PB138314BL - 1	PBW	-41	0.0259	-	7470A	PPB	12 Aug 2021 10:23:34	U	SMPL
PB138314BS - 1	LCSW	16085	3.9147	-	7470A	PPB	12 Aug 2021 10:25:51	U	SMPL
M3316-02 - 1	GROUND-WATER	3145	0.7942	-	7470A	PPB	12 Aug 2021 10:30:50	U	SMPL
CCV40 - 1	CCV40	20514	4.9827	-	7470A	PPB	12 Aug 2021 10:33:06	U	SMPL
CCB40 - 1	CCB40	-130	0.0044	-	7470A	PPB	12 Aug 2021 10:35:21	U	SMPL
M3316-02DUP - 1	GROUND-WATERDUP	2972	0.7525	-	7470A	PPB	12 Aug 2021 10:37:36	U	SMPL
M3316-02MS - 1	GROUND-WATERMS	18619	4.5258	-	7470A	PPB	12 Aug 2021 10:39:51	U	SMPL
M3316-02MSD - 1	GROUND-WATERMSD	18463	4.4881	-	7470A	PPB	12 Aug 2021 10:42:07	U	SMPL
PB138336BL - 1	PBW	-128	0.0049	-	7470A	PPB	12 Aug 2021 10:44:22	U	SMPL
PB138336BS - 1	LCSW	15359	3.7396	-	7470A	PPB	12 Aug 2021 10:46:38	U	SMPL
M3345-01 - 1	DRUM-1-3-18B	226	0.0903	-	7470A	PPB	12 Aug 2021 10:48:55	U	SMPL
M3346-02 - 1	LPA-1	323	0.1137	-	7470A	PPB	12 Aug 2021 10:51:11	U	SMPL
M3348-01 - 1	SPEEDIDRI-TRE-1237	717	0.2087	-	7470A	PPB	12 Aug 2021 10:58:26	U	SMPL
M3348-01DUP - 1	SPEEDIDRI-TRE-1237DUP	638	0.1896	-	7470A	PPB	12 Aug 2021 11:00:42	U	SMPL
M3348-01MS - 1	SPEEDIDRI-TRE-1237MS	16438	3.9998	-	7470A	PPB	12 Aug 2021 11:02:58	U	SMPL
CCV41 - 1	CCV41	20401	4.9555	-	7470A	PPB	12 Aug 2021 11:05:14	U	SMPL
CCB41 - 1	CCB41	-132	0.0040	-	7470A	PPB	12 Aug 2021 11:07:29	U	SMPL
M3348-01MSD - 1	SPEEDIDRI-TRE-1237MSD	16053	3.9070	-	7470A	PPB	12 Aug 2021 11:09:45	U	SMPL
PB138313TB - 1	PB138295TB	-70	0.0189	-	7470A	PPB	12 Aug 2021 11:12:00	U	SMPL
PB138336TB - 1	PB138318TB	138	0.0691	-	7470A	PPB	12 Aug 2021 11:18:48	U	SMPL
M3331-13LX5 - 1		-51	0.0235	-	7470A	PPB	12 Aug 2021 11:21:04	U	SMPL
M3331-13A - 1		15144	3.6878	-	7470A	PPB	12 Aug 2021 11:23:19	U	SMPL
M3316-02LX5 - 1		485	0.1527	-	7470A	PPB	12 Aug 2021 11:25:35	U	SMPL
M3316-02A - 1		18678	4.5400	-	7470A	PPB	12 Aug 2021 11:27:52	U	SMPL
M3348-01LX5 - 1		172	0.0773	-	7470A	PPB	12 Aug 2021 11:32:51	U	SMPL
M3348-01A - 1		17975	4.3705	-	7470A	PPB	12 Aug 2021 11:35:08	U	SMPL
CCV42 - 1	CCV42	20376	4.9495	-	7470A	PPB	12 Aug 2021 11:37:24	U	SMPL
CCB42 - 1	CCB42	-113	0.0085	-	7470A	PPB	12 Aug 2021 11:39:39	U	SMPL