



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 08/13/2021

OVEN IN TEMP1: 104 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 08/12/2021 16:00

DATE & TIME OUT1: 08/13/2021 08:00

OVEN IN TEMP2: 103 °C

OVEN OUT TEMP2: 103 °C

DATE & TIME IN2: 08/13/2021 08:30

DATE & TIME OUT2: 08/13/2021 10:00

Run Number: LB115802

BalanceID: Wc Sc-06

OvenID: Wc Oven-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB115802BL	LB115802BL	789.6377	789.6377	100	789.6378	789.6378	789.6378	0.0001	1
2	M3371-03	JET-GROUT	75.5471	75.5471	100	75.7092	75.7092	75.7092	0.1621	1621
3	M3371-03DUP	JET-GROUTDUP	68.8166	68.8166	100	68.9767	68.9767	68.9767	0.1601	1601

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

15802

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ASTM TS M3371

WorkList ID : 151346

Department : Wet-Chemistry

Date : 08-12-2021 08:49:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/24/2021	Solid	M3371-03	ASTM TS		TULL02	D11	JET-GROUT	08/10/2021	SM2540 B

Date/Time 08/12/21 12:40
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)

Date/Time 08/12/21 16:10
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)