

LB115809


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49111



A= 0.0000e+000

B= 2.0445e-004 *slope*C= -4.6181e-003 *y-intercept*

Rho= 0.9999751

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.006	0.006	50	0.000	50					11
0.05	0.050					573					15
0.20	0.200	0.230	0.030	1146	0.0 %	1146					1
2.50	2.500	2.474	-0.026	12124	0.0 %	12124					0
5.00	5.000	4.986	-0.014	24410	0.0 %	24410					0
7.50	7.500	7.469	-0.031	36553	0.0 %	36553					0
10.0	10.000	10.036	0.036	49111	0.0 %	49111					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	50	-		0 SFAM01.1	PPB	8/12/2021 11:58	Std
	0.2 S01	1146	-		0.2 SFAM01.1	PPB	8/12/2021 12:00	Std
	2.5 S02	12124	-		2.5 SFAM01.1	PPB	8/12/2021 12:02	Std
	5 S03	24410	-		5 SFAM01.1	PPB	8/12/2021 12:05	Std
	7.5 S04	36553	-		7.5 SFAM01.1	PPB	8/12/2021 12:07	Std
	10 S05	49111	-		10 SFAM01.1	PPB	8/12/2021 12:14	Std
ICV059	ICV059	18726	3.8239 -		SFAM01.1	PPB	8/12/2021 12:17	SMPL
ICB059	ICB059	-62	-0.0173 -		SFAM01.1	PPB	8/12/2021 12:19	SMPL
CCV043	CCV043	23889	4.8795 -		SFAM01.1	PPB	8/12/2021 12:22	SMPL
CCB043	CCB043	-48	-0.0144 -		SFAM01.1	PPB	8/12/2021 12:24	SMPL
PB138350BL	PBS350	21	-0.0003 -		SFAM01.1	PPB	8/12/2021 12:28	SMPL
M3233-01	BG2X1	73950	15.1143 -		SFAM01.1	PPB	8/12/2021 12:31	SMPL
M3233-02	BG2X2	2376	0.4812 -		SFAM01.1	PPB	8/12/2021 12:33	SMPL
M3233-03	BG2X2D	2620	0.531 -		SFAM01.1	PPB	8/12/2021 12:35	SMPL
M3233-04	BG2X2S	15737	3.2128 -		SFAM01.1	PPB	8/12/2021 12:37	SMPL
PB138353BL	PBS353	-22	-0.0091 -		SFAM01.1	PPB	8/12/2021 12:40	SMPL
M3243-01	JCE01	5301	1.0792 -		SFAM01.1	PPB	8/12/2021 12:42	SMPL
M3243-02	JCE02	20592	4.2054 -		SFAM01.1	PPB	8/12/2021 12:44	SMPL
M3243-03	JCE03	5334	1.0859 -		SFAM01.1	PPB	8/12/2021 12:47	SMPL
M3243-04	JCE04	21457	4.3822 -		SFAM01.1	PPB	8/12/2021 12:49	SMPL
M3243-05	JCE05	5258	1.0704 -		SFAM01.1	PPB	8/12/2021 12:51	SMPL
M3243-06	JCE05D	5621	1.1446 -		SFAM01.1	PPB	8/12/2021 12:53	SMPL
M3243-07	JCE05S	18324	3.7417 -		SFAM01.1	PPB	8/12/2021 12:56	SMPL
M3243-08	JCE06	8053	1.6418 -		SFAM01.1	PPB	8/12/2021 12:58	SMPL
M3243-09	JCE07	7071	1.441 -		SFAM01.1	PPB	8/12/2021 13:00	SMPL
M3243-10	JCE08	5888	1.1992 -		SFAM01.1	PPB	8/12/2021 13:02	SMPL
M3243-11	JCE11	399	0.077 -		SFAM01.1	PPB	8/12/2021 13:05	SMPL
M3243-12	JCE12	380	0.0731 -		SFAM01.1	PPB	8/12/2021 13:07	SMPL
PB138354BL	PBS354	25	0.0005 -		SFAM01.1	PPB	8/12/2021 13:09	SMPL
M3245-01	MC0AA0	614	0.1209 -		SFAM01.1	PPB	8/12/2021 13:11	SMPL
CCV044	CCV044	24208	4.9447 -		SFAM01.1	PPB	8/12/2021 13:14	SMPL
CCB044	CCB044	-36	-0.012 -		SFAM01.1	PPB	8/12/2021 13:16	SMPL
M3245-02	MC0AA1	2050	0.4145 -		SFAM01.1	PPB	8/12/2021 13:18	SMPL
M3245-03	MC0AA2	2467	0.4998 -		SFAM01.1	PPB	8/12/2021 13:21	SMPL
M3245-04	MC0AA3	1847	0.373 -		SFAM01.1	PPB	8/12/2021 13:23	SMPL
M3245-05	MC0AA4	4753	0.9671 -		SFAM01.1	PPB	8/12/2021 13:25	SMPL
M3245-06	MC0AA5	807	0.1604 -		SFAM01.1	PPB	8/12/2021 13:27	SMPL
M3245-07	MC0AA6	3204	0.6504 -		SFAM01.1	PPB	8/12/2021 13:30	SMPL
M3245-08	MC0AA6D	3103	0.6298 -		SFAM01.1	PPB	8/12/2021 13:37	SMPL
M3245-09	MC0AA6S	17205	3.5129 -		SFAM01.1	PPB	8/12/2021 13:39	SMPL
M3245-10	MC0AA7	1994	0.4031 -		SFAM01.1	PPB	8/12/2021 13:41	SMPL
M3245-11	MC0AA8	3094	0.6279 -		SFAM01.1	PPB	8/12/2021 13:43	SMPL
M3245-12	MC0AA9	708	0.1401 -		SFAM01.1	PPB	8/12/2021 13:46	SMPL
M3245-13	MC0AB0	2227	0.4507 -		SFAM01.1	PPB	8/12/2021 13:48	SMPL
M3245-14	MC0AB4	3023	0.6134 -		SFAM01.1	PPB	8/12/2021 13:50	SMPL
M3233-01DLX5	JCE01	15738	3.213 -		SFAM01.1	PPB	8/12/2021 13:52	SMPL

LB115809

INSTRUMENT ID : CV1

CCV045	CCV045	24325	4.9686 -	SFAM01.1	PPB	8/12/2021 13:55 SMPL
CCB045	CCB045	-32	-0.0112 -	SFAM01.1	PPB	8/12/2021 13:57 SMPL