



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115826

Slope : 98.4

BalanceID: WC SC-04

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.3	4.01	08/12/2021	17:15
2	CAL2	1	Water	NA	NA	22.3	7.00	08/12/2021	17:16
3	CAL3	1	Water	NA	NA	22.3	10.03	08/12/2021	17:19
4	ICV	1	Water	NA	NA	22.4	7.00	08/12/2021	17:20
5	CCV1	1	Water	NA	NA	22.4	2.01	08/12/2021	17:24
6	M3401-01	1	Solid	20.02	20	24.6	8.74	08/12/2021	17:25
7	M3401-01DUP	1	Solid	20.03	20	24.7	8.75	08/12/2021	17:26
8	CCV2	1	Water	NA	NA	22.4	12.02	08/12/2021	17:30

WORKLIST(Hardcopy Internal Chain)

VB115826

WorkList Name : PH M3401

WorkList ID : 151386

Department : Wet-Chemistry

Date : 08-12-2021 17:05:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/12/2021	Solid	M3401-01	pH	Cool 4 deg C	PSEG05	D51	BR-2-081221	08/12/2021	9045D

Date/Time

08/12/21 17:10

Raw Sample Received by:

JD (WC)

Raw Sample Relinquished by:

JD (WC)

Date/Time

08/12/21

Raw Sample Received by:

JD (WC)

Raw Sample Relinquished by:

JD (WC)