



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115827

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.3	4.01	08/12/2021	17:15
2	CAL2	1	Water	NA	NA	22.3	7.00	08/12/2021	17:16
3	CAL3	1	Water	NA	NA	22.3	10.03	08/12/2021	17:19
4	ICV	1	Water	NA	NA	22.4	7.00	08/12/2021	17:20
5	CCV1	1	Water	NA	NA	22.4	2.01	08/12/2021	17:24
6	M3391-01	1	Water	NA	NA	22.6	8.79	08/12/2021	17:35
7	M3391-01DUP	1	Water	NA	NA	22.7	8.80	08/12/2021	17:36
8	CCV2	1	Water	NA	NA	22.4	12.02	08/12/2021	17:38

23115827

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3391 WorkList ID : 151383 Department : Wet-Chemistry Date : 08-12-2021 16:56:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/12/2021	Water	M3391-01	H Corrosivity	Cool 4 deg C	PSEG03	D51	FRAC-TANK-F06031	08/12/2021	9040C

Date/Time 08.12.21 17:00
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: [Signature]

Date/Time 08.12.21 18:20
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: JP (WC)