

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: apatel

Parameter: ASTM COD

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB115848

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP92453
COD calibration std. 100 ppm	WP92450
COD calibration std. 10 ppm	WP92448
COD calibration std. 150 ppm	WP92451
COD calibration std. 50 ppm	WP92449
COD calibration std. 0 ppm	WP92447
COD CCV std, 50ppm	WP92452
COD Digestion Vials Low Level 0-150Mg/L	W2824

Temp In (C): <u>150</u>	Date In: <u>08/13/2021</u>	Time In: <u>09:05</u>
Temp Out (C): <u>149</u>	Date Out: <u>08/13/2021</u>	Time Out: <u>11:06</u>

Intercept: -0.0239

Slope: 0.9843

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	1.000	1.04		08/09/2021	12:31
2	CAL2	10	1	Water	9.000	9.168		08/09/2021	12:31
3	CAL3	50	1	Water	49.000	49.806		08/09/2021	12:32
4	CAL4	100	1	Water	98.000	99.587		08/09/2021	12:32
5	CAL5	150	1	Water	148.000	150.385		08/09/2021	12:33

Analysis Method: SM5220 DANALYST: apatelParameter: ASTM CODSUPERVISOR REVIEW BY: ketankumarRun Number: LB115848

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	48.000	48.790	08/09/2021	12:36
2	ICB		NA	NA	1	Water	1.000	1.040	08/09/2021	12:36
3	CCV1	50	NA	NA	1	Water	50.000	50.822	08/16/2021	09:27
4	CCB1		NA	NA	1	Water	0.000	0.024	08/16/2021	09:27
5	LB115848BL		NA	NA	1	Water	1.000	1.040	08/16/2021	09:28
6	LB115848BS	50	NA	NA	1	Water	49.000	49.806	08/16/2021	09:28
7	M3371-03		NA	NA	1	Water	14.000	14.248	08/16/2021	09:30
8	M3371-03DUP		NA	NA	1	Water	15.000	15.264	08/16/2021	09:30
9	M3371-03MS	50	NA	NA	1	Water	65.000	66.061	08/16/2021	09:31
10	M3371-03MSD	50	NA	NA	1	Water	65.000	66.061	08/16/2021	09:31
11	CCV2	50	NA	NA	1	Water	49.000	49.806	08/16/2021	09:32
12	CCB2		NA	NA	1	Water	0.000	0.024	08/16/2021	09:32