



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115857

Slope : 99.2

BalanceID: WC SC-04

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.6	4.01	08/14/2021	11:15
2	CAL2	1	Water	NA	NA	22.6	7.01	08/14/2021	11:16
3	CAL3	1	Water	NA	NA	22.6	10.03	08/14/2021	11:17
4	ICV	1	Water	NA	NA	22.6	6.99	08/14/2021	11:20
5	CCV1	1	Water	NA	NA	22.6	2.01	08/14/2021	11:25
6	M3424-02	1	Solid	20.03	20	24.0	9.85	08/14/2021	11:30
7	M3424-05	1	Solid	20.02	20	23.7	10.16	08/14/2021	11:32
8	M3424-08	1	Solid	20.03	20	23.5	9.12	08/14/2021	11:35
9	M3424-11	1	Solid	20.04	20	23.8	9.05	08/14/2021	11:37
10	M3424-14	1	Solid	20.02	20	23.7	10.30	08/14/2021	11:40
11	M3425-02	1	Solid	20.03	20	23.7	8.86	08/14/2021	11:42
12	M3425-05	1	Solid	20.02	20	24.2	8.65	08/14/2021	11:45
13	M3425-08	1	Solid	20.03	20	24.3	9.48	08/14/2021	11:48
14	M3425-11	1	Solid	20.02	20	24.1	10.27	08/14/2021	11:50
15	M3425-14	1	Solid	20.03	20	23.9	9.72	08/14/2021	12:00
16	CCV2	1	Water	NA	NA	22.5	12.02	08/14/2021	12:03
17	M3425-17	1	Solid	20.05	20	24.1	10.81	08/14/2021	12:06
18	M3425-20	1	Solid	20.03	20	24.2	11.21	08/14/2021	12:09
19	M3425-20DUP	1	Solid	20.04	20	24.3	11.22	08/14/2021	12:10
20	CCV3	1	Water	NA	NA	22.5	2.01	08/14/2021	12:22

VR 115857

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3424 WorkList ID : 151446 Department : Wet-Chemistry Date : 08-14-2021 11:03:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/13/2021	Solid	M3424-02	pH	Cool 4 deg C	JPCL01	D52	18N	08/13/2021	9045D
08/13/2021	Solid	M3424-05	pH	Cool 4 deg C	JPCL01	D52	18S	08/13/2021	9045D
08/13/2021	Solid	M3424-08	pH	Cool 4 deg C	JPCL01	D52	19N	08/13/2021	9045D
08/13/2021	Solid	M3424-11	pH	Cool 4 deg C	JPCL01	D52	19S	08/13/2021	9045D
08/13/2021	Solid	M3424-14	pH	Cool 4 deg C	JPCL01	D52	20N	08/13/2021	9045D
08/13/2021	Solid	M3425-02	pH	Cool 4 deg C	JPCL01	D31	20S	08/13/2021	9045D
08/13/2021	Solid	M3425-05	pH	Cool 4 deg C	JPCL01	D31	21N	08/13/2021	9045D
08/13/2021	Solid	M3425-08	pH	Cool 4 deg C	JPCL01	D31	21S	08/13/2021	9045D
08/13/2021	Solid	M3425-11	pH	Cool 4 deg C	JPCL01	D31	22N	08/13/2021	9045D
08/13/2021	Solid	M3425-14	pH	Cool 4 deg C	JPCL01	D31	22S	08/13/2021	9045D
08/13/2021	Solid	M3425-17	pH	Cool 4 deg C	JPCL01	D31	23N	08/13/2021	9045D
08/13/2021	Solid	M3425-20	pH	Cool 4 deg C	JPCL01	D31	23S	08/13/2021	9045D

Date/Time 08-14-21 11:10
 Raw Sample Received by: SR JWC
 Raw Sample Relinquished by: J.C. (sw)

Date/Time 08-14-21 16:20
 Raw Sample Received by: J.C. (sw)
 Raw Sample Relinquished by: J.C. (sw)