



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115876

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	08/16/2021	09:05
2	M3360-01		1	9	100.00	>212.0	>212.0	08/16/2021	09:35
3	M3360-01DUP		1	9	100.00	>212.0	>212.0	08/16/2021	09:58
4	M3360-02		1	10	100.00	>212.0	>212.0	08/16/2021	10:25
5	M3360-03		1	10	100.00	>212.0	>212.0	08/16/2021	10:50
6	M3360-04		1	9	100.00	>212.0	>212.0	08/16/2021	11:15
7	M3360-05		1	10	100.00	>212.0	>212.0	08/16/2021	11:40
8	M3361-14		1	12	100.00	>212.0	>212.0	08/16/2021	12:10
9	M3361-15		1	10	100.00	>212.0	>212.0	08/16/2021	12:35
10	M3361-16		1	11	100.00	>212.0	>212.0	08/16/2021	12:55
11	M3361-17		1	10	100.00	>212.0	>212.0	08/16/2021	13:20
12	M3391-01		1	12	100.00	>212.0	>212.0	08/16/2021	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp81321 WorkList ID : 151401 Department : Wet-Chemistry Date : 08-13-2021 09:43:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/18/2021	Water	M3360-01	Flash Point	Cool 4 deg C	PSEG03	D42	41266	08/11/2021	1010B
08/18/2021	Water	M3360-02	Flash Point	Cool 4 deg C	PSEG03	D42	41265	08/11/2021	1010B
08/18/2021	Water	M3360-03	Flash Point	Cool 4 deg C	PSEG03	D42	41920	08/11/2021	1010B
08/18/2021	Water	M3360-04	Flash Point	Cool 4 deg C	PSEG03	D42	41935	08/11/2021	1010B
08/18/2021	Water	M3360-05	Flash Point	Cool 4 deg C	PSEG03	D42	31109	08/11/2021	1010B
08/18/2021	Water	M3361-14	Flash Point	Cool 4 deg C	PSEG03	D42	41909	08/11/2021	1010B
08/18/2021	Water	M3361-15	Flash Point	Cool 4 deg C	PSEG03	D42	07-08-11	08/11/2021	1010B
08/18/2021	Water	M3361-16	Flash Point	Cool 4 deg C	PSEG03	D42	12-42-49	08/11/2021	1010B
08/18/2021	Water	M3361-17	Flash Point	Cool 4 deg C	PSEG03	D42	40-43	08/11/2021	1010B
08/19/2021	Water	M3391-01	Flash Point	Cool 4 deg C	PSEG03	D51	FRAC-TANK-F06031	08/12/2021	1010B

Date/Time 08/16/21 08:30
 Raw Sample Received by: KS (WOC)
 Raw Sample Relinquished by: RJ CEXF-100b)

Date/Time 08/16/21 13:50
 Raw Sample Received by: RJ CEXF-100b)
 Raw Sample Relinquished by: KS (WOC)