



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115877

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	08/16/2021	09:05
2	M3361-05		1	10	100.00	>212.0	>212.0	08/16/2021	14:15
3	M3361-05DUP		1	10	100.00	>212.0	>212.0	08/16/2021	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 115877

WorkList Name : fp81321-1 WorkList ID : 151402 Department : Wet-Chemistry Date : 08-13-2021 09:44:29

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/18/2021	Solid	M3361-05 ✓	Flash Point		Cool 4 deg C	PSEG03	D42	41910	08/11/2021	1010A



Date/Time 08/16/21 13:50
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JD (WC)

Date/Time 08/16/21 14:45
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)