



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115887

Slope : 98.7

BalanceID: WC SC-04

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.3	4.01	08/16/2021	16:12
2	CAL2	1	Water	NA	NA	22.3	7.00	08/16/2021	16:13
3	CAL3	1	Water	NA	NA	22.3	10.04	08/16/2021	16:14
4	ICV	1	Water	NA	NA	22.3	7.01	08/16/2021	16:15
5	CCV1	1	Water	NA	NA	22.3	2.01	08/16/2021	16:16
6	M3428-03	1	Solid	20.03	20	23.2	6.89	08/16/2021	16:19
7	M3428-03DUP	1	Solid	20.04	20	23.3	6.90	08/16/2021	16:20
8	M3428-06	1	Solid	20.02	20	22.3	6.10	08/16/2021	16:22
9	M3428-09	1	Solid	20.03	20	23.1	6.99	08/16/2021	16:25
10	M3428-12	1	Solid	20.01	20	23.2	6.62	08/16/2021	16:30
11	M3428-15	1	Solid	20.02	20	22.7	7.83	08/16/2021	16:32
12	M3428-18	1	Solid	20.03	20	22.6	7.57	08/16/2021	16:35
13	M3441-01	1	Solid	20.05	20	23.5	7.83	08/16/2021	16:40
14	M3441-04	1	Solid	20.04	20	23.7	8.29	08/16/2021	16:42
15	CCV2	1	Water	NA	NA	22.1	12.02	08/16/2021	16:45

115887

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3341 WorkList ID : 151487 Department : Wet-Chemistry Date : 08-16-2021 14:26:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/11/2021	Solid	M3428-03	Corrosivity	Cool 4 deg C	LIRO01	D52	SB-01-COMP	08/11/2021	9045D
08/11/2021	Solid	M3428-06	Corrosivity	Cool 4 deg C	LIRO01	D52	SB-02-COMP	08/11/2021	9045D
08/11/2021	Solid	M3428-09	Corrosivity	Cool 4 deg C	LIRO01	D52	SB-03-COMP	08/11/2021	9045D
08/12/2021	Solid	M3428-12	Corrosivity	Cool 4 deg C	LIRO01	D52	SB-04-COMP	08/12/2021	9045D
08/12/2021	Solid	M3428-15	Corrosivity	Cool 4 deg C	LIRO01	D52	SB-05-COMP	08/12/2021	9045D
08/12/2021	Solid	M3428-18	Corrosivity	Cool 4 deg C	LIRO01	D52	SB-06-COMP	08/12/2021	9045D
08/16/2021	Solid	M3441-01	Corrosivity	Cool 4 deg C	PSEG03	D61	WC-4	08/16/2021	9045D
08/16/2021	Solid	M3441-04	Corrosivity	Cool 4 deg C	PSEG03	D61	WC-5	08/16/2021	9045D

Date/Time 08-16-21 14:30
 Raw Sample Received by: JP (wgc)
 Raw Sample Relinquished by: JP (wgc)

Date/Time 08-16-21 16:30
 Raw Sample Received by: JP (wgc)
 Raw Sample Relinquished by: JP (wgc)