

LB115899


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

53773



A= 0.0000e+000

B= 1.8570e-004

C= 2.2366e-002

Rho= 0.9999712

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.030	0.030	41	0.000	41					
0.2	0.200	0.192	-0.008	912	0.0 %	912					-4
2.5	2.500	2.500	-0.000	13340	0.0 %	13340					0
5.0	5.000	4.945	-0.055	26508	0.0 %	26508					-1
7.5	7.500	7.525	0.025	40403	0.0 %	40403					0
10.0	10.000	10.008	0.008	53773	0.0 %	53773					0

LB115899 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 17 Aug 2021 14:24:48					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7471B	PPB	17 Aug 2021 14:25:56	S	Std
0.2 - 1	50.2	912	-	0.2000	7471B	PPB	17 Aug 2021 14:28:13	S	Std
2.5 - 1	52.5	13340	-	2.5000	7471B	PPB	17 Aug 2021 14:30:30	S	Std
5.0 - 1	55	26508	-	5.0000	7471B	PPB	17 Aug 2021 14:32:47	S	Std
7.5 - 1	57.5	40403	-	7.5000	7471B	PPB	17 Aug 2021 14:35:03	S	Std
10.0 - 1	51.0	53773	-	10.0000	7471B	PPB	17 Aug 2021 14:37:20	S	Std
ICV57 - 1	ICV57	21271	3.9725	-	7471B	PPB	17 Aug 2021 14:40:06	U	SMPL
ICB57 - 1	ICB57	-83	0.0070	-	7471B	PPB	17 Aug 2021 14:42:20	U	SMPL
CCV74 - 1	CCV74	26961	5.0291	-	7471B	PPB	17 Aug 2021 14:44:35	U	SMPL
CCB74 - 1	CCB74	-88	0.0060	-	7471B	PPB	17 Aug 2021 14:46:50	U	SMPL
CRA - 1	CRA	896	0.1888	-	7471B	PPB	17 Aug 2021 14:49:05	U	SMPL
HighStd - 1	HighStd	54024	10.0549	-	7471B	PPB	17 Aug 2021 14:51:20	U	SMPL
ChkStd - 1	ChkStd	38076	7.0932	-	7471B	PPB	17 Aug 2021 14:53:39	U	SMPL
PB138507BL - 1	PBS	-116	0.0008	-	7471B	PPB	17 Aug 2021 14:55:55	U	SMPL
PB138507BS - 1	LCSS	22508	4.2022	-	7471B	PPB	17 Aug 2021 14:58:12	U	SMPL
M3421-02 - 1	340	3446	0.6623	-	7471B	PPB	17 Aug 2021 15:00:28	U	SMPL
M3421-04 - 1	72-11966	20401	3.8109	-	7471B	PPB	17 Aug 2021 15:02:45	U	SMPL
M3424-02 - 1	18N	23272	4.3441	-	7471B	PPB	17 Aug 2021 15:05:02	U	SMPL
M3424-05 - 1	18S	14067	2.6347	-	7471B	PPB	17 Aug 2021 15:07:21	U	SMPL
M3424-08 - 1	19N	18308	3.4222	-	7471B	PPB	17 Aug 2021 15:09:36	U	SMPL
CCV75 - 1	CCV75	26908	5.0193	-	7471B	PPB	17 Aug 2021 15:11:51	U	SMPL
CCB75 - 1	CCB75	-98	0.0042	-	7471B	PPB	17 Aug 2021 15:14:06	U	SMPL
M3424-11 - 1	19S	28992	5.4063	-	7471B	PPB	17 Aug 2021 15:16:22	U	SMPL
M3424-14 - 1	20N	14880	2.7856	-	7471B	PPB	17 Aug 2021 15:18:38	U	SMPL
M3425-02 - 1	20S	43847	8.1649	-	7471B	PPB	17 Aug 2021 15:20:54	U	SMPL
M3425-05 - 1	21N	8592	1.6179	-	7471B	PPB	17 Aug 2021 15:23:09	U	SMPL
M3425-08 - 1	21S	11994	2.2497	-	7471B	PPB	17 Aug 2021 15:25:26	U	SMPL
M3425-11 - 1	22N	44068	8.2060	-	7471B	PPB	17 Aug 2021 15:27:42	U	SMPL
M3425-14 - 1	22S	7364	1.3899	-	7471B	PPB	17 Aug 2021 15:29:58	U	SMPL
M3425-17 - 1	23N	36336	6.7701	-	7471B	PPB	17 Aug 2021 15:32:16	U	SMPL
M3425-20 - 1	23S	65812	12.2439	-	7471B	PPB	17 Aug 2021 15:34:33	U	SMPL
M3428-02 - 1	SB-01-COMP	318	0.0814	-	7471B	PPB	17 Aug 2021 15:36:50	U	SMPL
CCV76 - 1	CCV76	27362	5.1036	-	7471B	PPB	17 Aug 2021 15:39:07	U	SMPL
CCB76 - 1	CCB76	-94	0.0049	-	7471B	PPB	17 Aug 2021 15:41:22	U	SMPL
M3428-05 - 1	SB-02-COMP	328	0.0833	-	7471B	PPB	17 Aug 2021 15:43:38	U	SMPL
M3428-08 - 1	SB-03-COMP	957	0.2001	-	7471B	PPB	17 Aug 2021 15:45:53	U	SMPL
M3428-11 - 1	SB-04-COMP	12129	2.2748	-	7471B	PPB	17 Aug 2021 15:48:08	U	SMPL
M3428-14 - 1	SB-05-COMP	2682	0.5204	-	7471B	PPB	17 Aug 2021 15:50:24	U	SMPL
M3428-17 - 1	SB-06-COMP	670	0.1468	-	7471B	PPB	17 Aug 2021 15:52:40	U	SMPL
M3428-17DUP - 1	SB-06-COMPDUP	811	0.1730	-	7471B	PPB	17 Aug 2021 15:54:56	U	SMPL
M3428-17MS - 1	SB-06-COMPMS	23399	4.3677	-	7471B	PPB	17 Aug 2021 15:57:12	U	SMPL
M3428-17MSD - 1	SB-06-COMPMSD	20988	3.9199	-	7471B	PPB	17 Aug 2021 15:59:29	U	SMPL
PB138508BL - 1	PBS	-43	0.0144	-	7471B	PPB	17 Aug 2021 16:01:46	U	SMPL
PB138508BS - 1	LCSS	22211	4.1470	-	7471B	PPB	17 Aug 2021 16:04:01	U	SMPL
CCV77 - 1	CCV77	26834	5.0056	-	7471B	PPB	17 Aug 2021 16:06:17	U	SMPL
CCB77 - 1	CCB77	-96	0.0045	-	7471B	PPB	17 Aug 2021 16:08:32	U	SMPL
M3412-01 - 1	PG-1	4088	0.7815	-	7471B	PPB	17 Aug 2021 16:10:48	U	SMPL
M3412-02 - 1	PG-2	3241	0.6242	-	7471B	PPB	17 Aug 2021 16:13:03	U	SMPL
M3412-03 - 1	PG-3	6391	1.2092	-	7471B	PPB	17 Aug 2021 16:15:18	U	SMPL
M3429-01 - 1	WC-1	8911	1.6772	-	7471B	PPB	17 Aug 2021 16:17:34	U	SMPL
M3429-04 - 1	WC-2	17020	3.1831	-	7471B	PPB	17 Aug 2021 16:19:49	U	SMPL
M3429-07 - 1	WC-3	23720	4.4273	-	7471B	PPB	17 Aug 2021 16:22:05	U	SMPL
M3431-01 - 1	1	690	0.1505	-	7471B	PPB	17 Aug 2021 16:24:22	U	SMPL
M3431-02 - 1	2	659	0.1447	-	7471B	PPB	17 Aug 2021 16:26:39	U	SMPL
M3431-03 - 1	3	701	0.1525	-	7471B	PPB	17 Aug 2021 16:28:56	U	SMPL
M3431-04 - 1	4	696	0.1516	-	7471B	PPB	17 Aug 2021 16:31:12	U	SMPL
CCV78 - 1	CCV78	26695	4.9797	-	7471B	PPB	17 Aug 2021 16:33:27	U	SMPL
CCB78 - 1	CCB78	-95	0.0047	-	7471B	PPB	17 Aug 2021 16:35:42	U	SMPL
M3431-05 - 1	5	701	0.1525	-	7471B	PPB	17 Aug 2021 16:37:58	U	SMPL
M3431-06 - 1	6	668	0.1464	-	7471B	PPB	17 Aug 2021 16:40:13	U	SMPL
M3431-07 - 1	7	656	0.1442	-	7471B	PPB	17 Aug 2021 16:42:29	U	SMPL
M3431-07DUP - 1	7DUP	655	0.1440	-	7471B	PPB	17 Aug 2021 16:44:44	U	SMPL
M3431-07MS - 1	7MS	20686	3.8638	-	7471B	PPB	17 Aug 2021 16:47:00	U	SMPL
M3431-07MSD - 1	7MSD	21465	4.0085	-	7471B	PPB	17 Aug 2021 16:49:16	U	SMPL
M3428-17LX5 - 1		52	0.0320	-	7471B	PPB	17 Aug 2021 16:51:32	U	SMPL
M3428-17A - 1		22404	4.1829	-	7471B	PPB	17 Aug 2021 16:53:49	U	SMPL
M3431-07LX5 - 1		31	0.0281	-	7471B	PPB	17 Aug 2021 16:56:06	U	SMPL
M3431-07A - 1		21451	4.0059	-	7471B	PPB	17 Aug 2021 16:58:22	U	SMPL
CCV79 - 1	CCV79	27135	5.0615	-	7471B	PPB	17 Aug 2021 17:00:38	U	SMPL
CCB79 - 1	CCB79	-124	-0.0007	-	7471B	PPB	17 Aug 2021 17:02:54	U	SMPL
M3425-20DLX2 - 1		32383	6.0360	-	7471B	PPB	17 Aug 2021 17:05:09	U	SMPL
CCV80 - 1	CCV80	26570	4.9565	-	7471B	PPB	17 Aug 2021 17:07:25	U	SMPL
CCB80 - 1	CCB80	-131	-0.0020	-	7471B	PPB	17 Aug 2021 17:09:41	U	SMPL