



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115902

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	08/17/2021	13:50
2	M3440-05		1	10	100.00	>212.0	>212.0	08/17/2021	14:15
3	M3440-05DUP		1	10	100.00	>212.0	>212.0	08/17/2021	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp81721 WorkList ID : 151507 Department : Wet-Chemistry Date : 08-17-2021 08:56:50

Due Date	Matrix	Sample	Test	Flash Point	Cool 4 deg C	PSEG03	D61	Raw Sample Storage Location	Customer	Customer Sample	Collect Date	Method
08/23/2021	Water	M3440-05	C	Flash Point	Cool 4 deg C	PSEG03	D61	DRUM-D1992			08/16/2021	1010B

Date/Time 08/17/21 12:25
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ CEXT (lab)

Date/Time 08/17/21 14:45
 Raw Sample Received by: RJ CEXT (lab)
 Raw Sample Relinquished by: KS (WC)