



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 8/18/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 08/17/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 08/18/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115904

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3445-02	TT-087-RW7-IDWSO-20210816-1	1	1.15	9.95	7.96	77.4	
M3445-03	TT-087-RW7-IDWSO-20210816-2	2	1.16	9.59	6.79	66.8	
M3445-04	TT-087-RW7-IDWSO-20210816-3	3	1.16	9.86	7.27	70.2	
M3447-01	TR-05-081721	4	1.17	9.66	9.04	92.7	
M3447-02	TR-05-081721-EPH-2	5	1.15	9.78	9.07	91.8	
M3449-01	TP1	6	1.12	9.6	8.96	92.5	
M3449-03	TP1-EPH-2	7	1.17	9.8	9.11	92.0	
M3449-04	TP2	8	1.17	9.91	8.79	87.2	
M3449-06	TP2-EPH-2	9	1.17	9.79	8.75	87.9	
M3449-07	TP3	10	1.15	9.79	8.8	88.5	
M3449-09	TP3-EPH-2	11	1.15	9.96	9.14	90.7	
M3449-10	TP4	12	1.18	9.76	8.73	88.0	
M3449-12	TP4-EPH-2	13	1.17	9.93	8.75	86.5	
M3449-13	TP5	14	1.15	9.95	9.22	91.7	
M3449-15	TP5-EPH-2	15	1.18	9.76	9.23	93.8	
M3449-16	TP6	16	1.19	9.92	8.63	85.2	
M3449-18	TP6-EPH-2	17	1.13	9.53	8.18	83.9	
M3449-19	CON-1	18	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3449-21	CON-2	19	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3449-23	CON-3	20	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3449-25	CON-4	21	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3449-27	CON-5	22	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3449-29	CON-6	23	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3450-01	P001-SS183-0001-01	24	1.16	9.76	9.51	97.1	
M3450-02	P001-SS183-0001-02	25	1.1	9.66	9.37	96.6	
M3450-03	P001-SS184-0001-01	26	1.15	9.95	9.6	96.0	
M3450-04	P001-SS185-0001-01	27	1.17	9.9	9.11	91.0	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 8/18/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 08/17/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 08/18/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115904

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3451-01	TP-0-5-COMP	28	1.14	9.77	8.6	86.4	
M3451-03	TP-0-5-COMP-EPH-2	29	1.14	9.55	8.33	85.5	
M3451-04	TP-5-10-COMP	30	1.12	9.67	8.96	91.7	
M3451-04D	TP-5-10-COMPDUP	31	1.14	9.69	8.99	91.8	
M3451-06	TP-5-10-COMP-EPH-2	32	1.17	9.55	8.89	92.1	
M3452-01	351	33	1.00	2.00	2.00	100.0	CONCRETE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$