

LB115907


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

42231



A= 0.0000e+000

B= 2.3725e-004

C= 4.5668e-003

Rho= 0.9997208

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.007	-0.007	-49	0.000	-49					18
0.2	0.200	0.236	0.036	975	0.0 %	975					-4
2.5	2.500	2.404	-0.096	10112	0.0 %	10112					3
5.0	5.000	5.151	0.151	21693	0.0 %	21693					-1
7.5	7.500	7.393	-0.107	31140	0.0 %	31140					0
10.0	10.000	10.024	0.024	42231	0.0 %	42231					

LB115907 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 18 Aug 2021 07:52:29					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-49	-	0.0000	7470A	PPB	18 Aug 2021 07:56:08	S	Std
0.2 - 1	50.2	975	-	0.2000	7470A	PPB	18 Aug 2021 07:58:25	S	Std
2.5 - 1	52.5	10112	-	2.5000	7470A	PPB	18 Aug 2021 08:00:42	S	Std
5.0 - 1	55	21693	-	5.0000	7470A	PPB	18 Aug 2021 08:02:59	S	Std
7.5 - 1	57.5	31140	-	7.5000	7470A	PPB	18 Aug 2021 08:05:15	S	Std
10.0 - 1	60	42231	-	10.0000	7470A	PPB	18 Aug 2021 08:10:12	S	Std
ICV59 - 1	ICV59	15336	3.6430	-	7470A	PPB	18 Aug 2021 08:12:56	U	SMPL
ICB59 - 1	ICB59	-85	-0.0156	-	7470A	PPB	18 Aug 2021 08:15:11	U	SMPL
CCV81 - 1	CCV81	20447	4.8556	-	7470A	PPB	18 Aug 2021 08:17:26	U	SMPL
CCB81 - 1	CCB81	-93	-0.0175	-	7470A	PPB	18 Aug 2021 08:19:41	U	SMPL
CRA - 1	CRA	966	0.2337	-	7470A	PPB	18 Aug 2021 08:21:55	U	SMPL
HighStd - 1	HighStd	41984	9.9652	-	7470A	PPB	18 Aug 2021 08:24:10	U	SMPL
ChkStd - 1	ChkStd	28601	6.7901	-	7470A	PPB	18 Aug 2021 08:26:26	U	SMPL
PB138510BL - 1	PBW	-99	-0.0189	-	7470A	PPB	18 Aug 2021 08:28:43	U	SMPL
PB138510BS - 1	LCSW	13889	3.2997	-	7470A	PPB	18 Aug 2021 08:30:59	U	SMPL
M3398-16 - 1	SWCS-04-0002	-24	-0.0011	-	7470A	PPB	18 Aug 2021 08:33:15	U	SMPL
M3398-19 - 1	SWCS-04-0205	44	0.0150	-	7470A	PPB	18 Aug 2021 08:35:31	U	SMPL
M3398-22 - 1	SWCS-04-0510	205	0.0532	-	7470A	PPB	18 Aug 2021 08:37:48	U	SMPL
M3419-02 - 1	SWCS-03-0002	90	0.0259	-	7470A	PPB	18 Aug 2021 08:40:05	U	SMPL
M3419-04 - 1	SWCS-03-0205	268	0.0681	-	7470A	PPB	18 Aug 2021 08:42:19	U	SMPL
CCV82 - 1	CCV82	20399	4.8442	-	7470A	PPB	18 Aug 2021 08:44:34	U	SMPL
CCB82 - 1	CCB82	-83	-0.0151	-	7470A	PPB	18 Aug 2021 08:46:50	U	SMPL
M3419-06 - 1	SWCS-03-0510	372	0.0928	-	7470A	PPB	18 Aug 2021 08:49:05	U	SMPL
M3419-08 - 1	SWCS-03-1015	6	0.0060	-	7470A	PPB	18 Aug 2021 08:51:20	U	SMPL
M3419-08DUP - 1	SWCS-03-1015DUP	9	0.0067	-	7470A	PPB	18 Aug 2021 08:53:35	U	SMPL
M3419-11 - 1	M3419-08MS	14983	3.5593	-	7470A	PPB	18 Aug 2021 08:55:51	U	SMPL
M3419-12 - 1	M3419-08MSD	14857	3.5294	-	7470A	PPB	18 Aug 2021 08:58:06	U	SMPL
M3419-14 - 1	SWCS-08-0002	84	0.0245	-	7470A	PPB	18 Aug 2021 09:00:23	U	SMPL
M3419-16 - 1	SWCS-08-0205	66	0.0202	-	7470A	PPB	18 Aug 2021 09:02:40	U	SMPL
M3419-18 - 1	SWCS-08-0510	0	0.0046	-	7470A	PPB	18 Aug 2021 09:04:56	U	SMPL
M3419-20 - 1	SWCS-08-1015	0	0.0046	-	7470A	PPB	18 Aug 2021 09:07:14	U	SMPL
M3419-22 - 1	SWCS-09-0002	123	0.0337	-	7470A	PPB	18 Aug 2021 09:09:29	U	SMPL
CCV83 - 1	CCV83	20441	4.8542	-	7470A	PPB	18 Aug 2021 09:11:45	U	SMPL
CCB83 - 1	CCB83	-62	-0.0101	-	7470A	PPB	18 Aug 2021 09:14:00	U	SMPL
M3419-24 - 1	SWCS-09-0205	204	0.0530	-	7470A	PPB	18 Aug 2021 09:16:15	U	SMPL
M3426-02 - 1	SWCS-09-0510	14	0.0079	-	7470A	PPB	18 Aug 2021 09:18:31	U	SMPL
M3426-04 - 1	2021-08-11-FD-02	18	0.0088	-	7470A	PPB	18 Aug 2021 09:20:46	U	SMPL
M3426-06 - 1	SWCS-09-1015	0	0.0046	-	7470A	PPB	18 Aug 2021 09:23:02	U	SMPL
M3426-08 - 1	SWCS-10-0002	124	0.0340	-	7470A	PPB	18 Aug 2021 09:25:17	U	SMPL
PB138511BL - 1	PBW	50	0.0164	-	7470A	PPB	18 Aug 2021 09:27:33	U	SMPL
PB138511BS - 1	LCSW	15544	3.6924	-	7470A	PPB	18 Aug 2021 09:29:49	U	SMPL
M3428-03 - 1	SB-01-COMP	-28	-0.0021	-	7470A	PPB	18 Aug 2021 09:32:06	U	SMPL
M3428-06 - 1	SB-02-COMP	-1	0.0043	-	7470A	PPB	18 Aug 2021 09:34:23	U	SMPL
M3428-09 - 1	SB-03-COMP	5	0.0058	-	7470A	PPB	18 Aug 2021 09:36:40	U	SMPL
CCV84 - 1	CCV84	20489	4.8656	-	7470A	PPB	18 Aug 2021 09:38:55	U	SMPL
CCB84 - 1	CCB84	-79	-0.0142	-	7470A	PPB	18 Aug 2021 09:41:11	U	SMPL
M3428-12 - 1	SB-04-COMP	13	0.0077	-	7470A	PPB	18 Aug 2021 09:43:26	U	SMPL
M3428-15 - 1	SB-05-COMP	9	0.0067	-	7470A	PPB	18 Aug 2021 09:45:41	U	SMPL
M3428-18 - 1	SB-06-COMP	7	0.0062	-	7470A	PPB	18 Aug 2021 09:47:57	U	SMPL
M3441-02 - 1	WC-4	7	0.0062	-	7470A	PPB	18 Aug 2021 09:50:12	U	SMPL
M3441-05 - 1	WC-5	15	0.0081	-	7470A	PPB	18 Aug 2021 09:52:27	U	SMPL
M3441-05DUP - 1	WC-5DUP	11	0.0072	-	7470A	PPB	18 Aug 2021 09:54:44	U	SMPL
M3441-05MS - 1	WC-5MS	14733	3.5000	-	7470A	PPB	18 Aug 2021 09:57:00	U	SMPL
M3441-05MSD - 1	WC-5MSD	14048	3.3374	-	7470A	PPB	18 Aug 2021 09:59:17	U	SMPL
PB138512BL - 1	PBW	-33	-0.0033	-	7470A	PPB	18 Aug 2021 10:01:33	U	SMPL
PB138512BS - 1	LCSW	15014	3.5666	-	7470A	PPB	18 Aug 2021 10:03:49	U	SMPL
CCV85 - 1	CCV85	20851	4.9514	-	7470A	PPB	18 Aug 2021 10:06:04	U	SMPL
CCB85 - 1	CCB85	-72	-0.0125	-	7470A	PPB	18 Aug 2021 10:08:20	U	SMPL
M3435-32 - 1	2021-08-10-EB-01	-51	-0.0075	-	7470A	PPB	18 Aug 2021 10:10:35	U	SMPL
M3435-32DUP - 1	2021-08-10-EB-01DUP	-14	0.0012	-	7470A	PPB	18 Aug 2021 10:12:51	U	SMPL
M3435-32MS - 1	2021-08-10-EB-01MS	15476	3.6762	-	7470A	PPB	18 Aug 2021 10:15:01	U	SMPL
M3435-32MSD - 1	2021-08-10-EB-01MSD	15367	3.6504	-	7470A	PPB	18 Aug 2021 10:17:17	U	SMPL
M3435-34 - 1	2021-08-11-EB-02	-106	-0.0206	-	7470A	PPB	18 Aug 2021 10:19:32	U	SMPL
M3435-36 - 1	2021-08-12-EB-03	-28	-0.0021	-	7470A	PPB	18 Aug 2021 10:21:48	U	SMPL
M3435-38 - 1	2021-08-13-EB-04	-21	-0.0004	-	7470A	PPB	18 Aug 2021 10:24:04	U	SMPL
PB138514BL - 1	PBW	5	0.0058	-	7470A	PPB	18 Aug 2021 10:26:21	U	SMPL
PB138514BS - 1	LCSW	13991	3.3239	-	7470A	PPB	18 Aug 2021 10:28:38	U	SMPL
M3391-01 - 1	FRAC-TANK-F06031	1291	0.3109	-	7470A	PPB	18 Aug 2021 11:00:54	U	SMPL
CCV86 - 1	CCV86	21044	4.9972	-	7470A	PPB	18 Aug 2021 11:03:09	U	SMPL
CCB86 - 1	CCB86	-72	-0.0125	-	7470A	PPB	18 Aug 2021 11:05:25	U	SMPL
M3391-01DUP - 1	FRAC-TANK-F06031DUP	1307	0.3147	-	7470A	PPB	18 Aug 2021 11:07:41	U	SMPL
M3391-01MS - 1	FRAC-TANK-F06031MS	18830	4.4720	-	7470A	PPB	18 Aug 2021 11:09:57	U	SMPL
M3391-01MSD - 1	FRAC-TANK-F06031MSD	17256	4.0985	-	7470A	PPB	18 Aug 2021 11:12:13	U	SMPL
M3435-31 - 1	2021-08-10-EB-01	-103	-0.0199	-	7470A	PPB	18 Aug 2021 11:14:30	U	SMPL
M3435-33 - 1	2021-08-11-EB-02	-35	-0.0037	-	7470A	PPB	18 Aug 2021 11:16:46	U	SMPL
M3435-35 - 1	2021-08-12-EB-03	-28	-0.0021	-	7470A	PPB	18 Aug 2021 11:19:02	U	SMPL
M3435-37 - 1	2021-08-13-EB-04	-15	0.0010	-	7470A	PPB	18 Aug 2021 11:21:19	U	SMPL
M3440-05 - 1	DRUM-D1992	121	0.0333	-	7470A	PPB	18 Aug 2021 11:23:36	U	SMPL
PB138510TB - 1	PB138474TB	23	0.0100	-	7470A	PPB	18 Aug 2021 11:25:53	U	SMPL
PB138511TB - 1	PB138477TB	10	0.0069	-	7470A	PPB	18 Aug 2021 11:28:10	U	SMPL
CCV87 - 1	CCV87	20320	4.8255	-	7470A	PPB	18 Aug 2021 11:30:26	U	SMPL

LB115907 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Aug 2021 07:52:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
CCB87 - 1	CCB87	-77	-0.0137		-7470A	PPB	18 Aug 2021 11:32:41	U	SMPL
PB138512TB - 1	PB138499TB	15	0.0081		-7470A	PPB	18 Aug 2021 11:34:57	U	SMPL
M3419-08LX5 - 1		39	0.0138		-7470A	PPB	18 Aug 2021 11:37:13	U	SMPL
M3419-08A - 1		15632	3.7132		-7470A	PPB	18 Aug 2021 11:39:29	U	SMPL
M3441-05LX5 - 1		-70	-0.0120		-7470A	PPB	18 Aug 2021 11:41:45	U	SMPL
M3441-05A - 1		15478	3.6767		-7470A	PPB	18 Aug 2021 11:44:02	U	SMPL
M3435-32LX5 - 1		-65	-0.0109		-7470A	PPB	18 Aug 2021 11:46:18	U	SMPL
M3435-32A - 1		15414	3.6615		-7470A	PPB	18 Aug 2021 11:48:34	U	SMPL
M3391-01LX5 - 1		221	0.0570		-7470A	PPB	18 Aug 2021 11:50:51	U	SMPL
M3391-01A - 1		18062	4.2898		-7470A	PPB	18 Aug 2021 11:53:08	U	SMPL
CCV88 - 1	CCV88	20968	4.9792		-7470A	PPB	18 Aug 2021 12:01:03	U	SMPL
CCB88 - 1	CCB88	-79	-0.0142		-7470A	PPB	18 Aug 2021 12:03:19	U	SMPL