



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115910

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	08/16/2021	15:20
2	CAL2	1	Water	NA	NA	21.1	7.00	08/16/2021	15:21
3	CAL3	1	Water	NA	NA	21.1	10.03	08/16/2021	15:22
4	ICV	1	Water	NA	NA	21.1	6.98	08/16/2021	15:25
5	CCV1	1	Water	NA	NA	21.1	2.01	08/16/2021	15:26
6	M3440-05	1	Water	NA	NA	22.1	6.80	08/16/2021	15:29
7	M3440-05DUP	1	Water	NA	NA	22.2	6.81	08/16/2021	15:30
8	CCV2	1	Water	NA	NA	21.2	12.02	08/16/2021	15:32