



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115923

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/18/2021	16:00
2	CAL2	1	Water	NA	NA	21.2	7.01	08/18/2021	16:01
3	CAL3	1	Water	NA	NA	21.2	10.03	08/18/2021	16:02
4	ICV	1	Water	NA	NA	21.2	6.99	08/18/2021	16:05
5	CCV1	1	Water	NA	NA	21.2	2.01	08/18/2021	16:10
6	M3475-01	1	Water	NA	NA	24.8	7.05	08/18/2021	16:13
7	M3475-01DUP	1	Water	NA	NA	24.9	7.06	08/18/2021	16:14
8	CCV2	1	Water	NA	NA	21.3	12.03	08/18/2021	16:15

VB115923

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3475 WorkList ID : 151562 Department : Wet-Chemistry Date : 08-18-2021 15:38:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/18/2021	Water	M3475-01	pH	Cool 4 deg C	PSEG03	E11	N5868-N34145	08/18/2021	9040C

15:40
Date/Time 08-18-21 15:40
Raw Sample Received by: YB (WC)
Raw Sample Relinquished by: CP SM

Date/Time 08-18-21 17:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JPTWC