


LB115931

7471B

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 1.8815e-004

C= 1.7446e-002

Rho= 0.9998905

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.023	0.023	32	0.000	32					-
0.2	0.200	0.195	-0.005	943	0.0 %	943					-3
2.5	2.500	2.413	-0.087	12732	0.0 %	12732					-3
5.0	5.000	5.079	0.079	26903	0.0 %	26903					2
7.5	7.500	7.536	0.036	39960	0.0 %	39960					0
10.0	10.000	9.953	-0.047	52809	0.0 %	52809					0

LB115931 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 19 Aug 2021 13:51:28					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	32	-	0.0000	7471B	PPB	19 Aug 2021 13:58:58	S	Std
0.2 - 1	50.2	943	-	0.2000	7471B	PPB	19 Aug 2021 14:01:15	S	Std
2.5 - 1	52.5	12732	-	2.5000	7471B	PPB	19 Aug 2021 14:03:33	S	Std
5.0 - 1	55	26903	-	5.0000	7471B	PPB	19 Aug 2021 14:05:49	S	Std
7.5 - 1	57.5	39960	-	7.5000	7471B	PPB	19 Aug 2021 14:08:06	S	Std
10.0 - 1	516	52809	-	10.0000	7471B	PPB	19 Aug 2021 14:15:46	S	Std
ICV62 - 1	ICV62	20745	3.9206	-	7471B	PPB	19 Aug 2021 14:18:27	U	SMPL
ICB62 - 1	ICB62	-16	0.0144	-	7471B	PPB	19 Aug 2021 14:20:41	U	SMPL
CCV08 - 1	CCV08	26831	5.0657	-	7471B	PPB	19 Aug 2021 14:22:56	U	SMPL
CCB08 - 1	CCB08	-12	0.0152	-	7471B	PPB	19 Aug 2021 14:25:11	U	SMPL
CRA - 1	CRA	947	0.1956	-	7471B	PPB	19 Aug 2021 14:27:26	U	SMPL
HighStd - 1	HighStd	53826	10.1448	-	7471B	PPB	19 Aug 2021 14:29:41	U	SMPL
ChkStd - 1	ChkStd	37908	7.1499	-	7471B	PPB	19 Aug 2021 14:31:56	U	SMPL
PB138582BL - 1	PBS	-130	-0.0070	-	7471B	PPB	19 Aug 2021 14:34:13	U	SMPL
PB138582BS - 1	LCSS	21989	4.1547	-	7471B	PPB	19 Aug 2021 14:36:30	U	SMPL
M3447-01 - 1	TR-05-081721	1758	0.3482	-	7471B	PPB	19 Aug 2021 14:38:46	U	SMPL
M3447-01DUP - 1	TR-05-081721DUP	1690	0.3354	-	7471B	PPB	19 Aug 2021 14:41:02	U	SMPL
M3447-01MS - 1	TR-05-081721MS	23884	4.5112	-	7471B	PPB	19 Aug 2021 14:43:19	U	SMPL
M3447-01MSD - 1	TR-05-081721MSD	20762	3.9238	-	7471B	PPB	19 Aug 2021 14:45:37	U	SMPL
M3449-01 - 1	TP1	1067	0.2182	-	7471B	PPB	19 Aug 2021 14:47:52	U	SMPL
CCV09 - 1	CCV09	27308	5.1555	-	7471B	PPB	19 Aug 2021 14:50:07	U	SMPL
CCB09 - 1	CCB09	-10	0.0156	-	7471B	PPB	19 Aug 2021 14:52:23	U	SMPL
M3449-04 - 1	TP2	1661	0.3300	-	7471B	PPB	19 Aug 2021 14:54:39	U	SMPL
M3449-07 - 1	TP3	3113	0.6032	-	7471B	PPB	19 Aug 2021 14:56:54	U	SMPL
M3449-10 - 1	TP4	4773	0.9155	-	7471B	PPB	19 Aug 2021 14:59:10	U	SMPL
M3449-13 - 1	TP5	3849	0.7416	-	7471B	PPB	19 Aug 2021 15:01:26	U	SMPL
M3449-16 - 1	TP6	15283	2.8929	-	7471B	PPB	19 Aug 2021 15:03:42	U	SMPL
M3449-19 - 1	CON-1	480	0.1078	-	7471B	PPB	19 Aug 2021 15:05:58	U	SMPL
M3449-21 - 1	CON-2	1622	0.3226	-	7471B	PPB	19 Aug 2021 15:08:15	U	SMPL
M3449-23 - 1	CON-3	4006	0.7712	-	7471B	PPB	19 Aug 2021 15:10:32	U	SMPL
M3449-25 - 1	CON-4	403	0.0933	-	7471B	PPB	19 Aug 2021 15:12:49	U	SMPL
M3449-27 - 1	CON-5	578	0.1262	-	7471B	PPB	19 Aug 2021 15:15:04	U	SMPL
CCV10 - 1	CCV10	27151	5.1259	-	7471B	PPB	19 Aug 2021 15:17:20	U	SMPL
CCB10 - 1	CCB10	1	0.0176	-	7471B	PPB	19 Aug 2021 15:19:36	U	SMPL
M3449-29 - 1	CON-6	364	0.0859	-	7471B	PPB	19 Aug 2021 15:21:52	U	SMPL
PB138583BL - 1	PBS	26	0.0223	-	7471B	PPB	19 Aug 2021 15:24:07	U	SMPL
PB138583BS - 1	LCSS	21169	4.0004	-	7471B	PPB	19 Aug 2021 15:26:23	U	SMPL
M3465-01 - 1	CON-A	181	0.0515	-	7471B	PPB	19 Aug 2021 15:28:38	U	SMPL
M3465-03 - 1	CON-B	851	0.1776	-	7471B	PPB	19 Aug 2021 15:30:54	U	SMPL
M3465-05 - 1	CON-C	255	0.0654	-	7471B	PPB	19 Aug 2021 15:33:10	U	SMPL
M3466-01 - 1	OK-03-081821	1570	0.3128	-	7471B	PPB	19 Aug 2021 15:35:27	U	SMPL
M3467-01 - 1	SAMPLE-11-(647.00)	4134	0.7953	-	7471B	PPB	19 Aug 2021 15:37:43	U	SMPL
M3467-03 - 1	SAMPLE-10-(410.75)	2799	0.5441	-	7471B	PPB	19 Aug 2021 15:40:01	U	SMPL
M3467-05 - 1	SAMPLE-9-(322.00)	3329	0.6438	-	7471B	PPB	19 Aug 2021 15:42:16	U	SMPL
CCV11 - 1	CCV11	27771	5.2426	-	7471B	PPB	19 Aug 2021 15:44:32	U	SMPL
CCB11 - 1	CCB11	-93	-0.0001	-	7471B	PPB	19 Aug 2021 15:46:47	U	SMPL
M3467-07 - 1	SAMPLE-8-(309.00)	10159	1.9289	-	7471B	PPB	19 Aug 2021 15:49:03	U	SMPL
M3467-09 - 1	SAMPLE-7-(291.00)	18544	3.5065	-	7471B	PPB	19 Aug 2021 15:51:20	U	SMPL
M3467-11 - 1	SAMPLE-1-(150.00)	25159	4.7511	-	7471B	PPB	19 Aug 2021 15:53:35	U	SMPL
M3467-13 - 1	SAMPLE-2-(161.00)	9040	1.7183	-	7471B	PPB	19 Aug 2021 15:55:51	U	SMPL
M3467-15 - 1	SAMPLE-3-(171.00)	4525	0.8688	-	7471B	PPB	19 Aug 2021 15:58:06	U	SMPL
M3467-17 - 1	SAMPLE-4-(177.00)	22882	4.3227	-	7471B	PPB	19 Aug 2021 16:00:22	U	SMPL
M3467-19 - 1	SAMPLE-5-(184.50)	13246	2.5097	-	7471B	PPB	19 Aug 2021 16:02:38	U	SMPL
M3467-21 - 1	SAMPLE-6-(203.30)	12931	2.4504	-	7471B	PPB	19 Aug 2021 16:04:56	U	SMPL
M3476-01 - 1	OR-03-081821	1271	0.2566	-	7471B	PPB	19 Aug 2021 16:07:12	U	SMPL
M3476-01DUP - 1	OR-03-081821DUP	1277	0.2577	-	7471B	PPB	19 Aug 2021 16:09:28	U	SMPL
CCV12 - 1	CCV12	27913	5.2693	-	7471B	PPB	19 Aug 2021 16:11:43	U	SMPL
CCB12 - 1	CCB12	-114	-0.0040	-	7471B	PPB	19 Aug 2021 16:13:58	U	SMPL
M3476-01MS - 1	OR-03-081821MS	20712	3.9144	-	7471B	PPB	19 Aug 2021 16:16:14	U	SMPL
M3476-01MSD - 1	OR-03-081821MSD	20963	3.9616	-	7471B	PPB	19 Aug 2021 16:18:30	U	SMPL
M3447-01LX5 - 1		-94	-0.0002	-	7471B	PPB	19 Aug 2021 16:20:45	U	SMPL
M3447-01A - 1		22154	4.1857	-	7471B	PPB	19 Aug 2021 16:23:01	U	SMPL
M3476-01LX5 - 1		141	0.0440	-	7471B	PPB	19 Aug 2021 16:25:17	U	SMPL
M3476-01A - 1		23125	4.3684	-	7471B	PPB	19 Aug 2021 16:27:35	U	SMPL
CCV13 - 1	CCV13	27495	5.1906	-	7471B	PPB	19 Aug 2021 16:29:51	U	SMPL
CCB13 - 1	CCB13	-56	0.0069	-	7471B	PPB	19 Aug 2021 16:32:08	U	SMPL