



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115932

Slope : 98.3

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2810         |
| buffer solution pH 7 yellow    | W2813         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2809         |
| buffer solution pH 7 yellow    | W2719         |
| Buffer Solution, PH2 (500ml)   | W2806         |
| Buffer Solution, PH12 (500ml)  | W2807         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 21.1             | 4.01        | 08/19/2021 | 14:00     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 21.1             | 7.00        | 08/19/2021 | 14:01     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 21.1             | 10.03       | 08/19/2021 | 14:02     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 21.2             | 6.99        | 08/19/2021 | 14:05     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 21.2             | 2.01        | 08/19/2021 | 14:06     |
| 6   | M3484-01    | 1  | Water  | NA          | NA          | 21.3             | 7.80        | 08/19/2021 | 14:10     |
| 7   | M3484-01DUP | 1  | Water  | NA          | NA          | 21.4             | 7.81        | 08/19/2021 | 14:11     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 21.1             | 12.02       | 08/19/2021 | 14:15     |

HP 115932

WORKLIST(Hardcopy Internal Chain)

WorkList Name : phm3484      WorkList ID : 151603      Department : Wet-Chemistry      Date : 08-19-2021 13:46:50

| Due Date   | Matrix | Sample   | Test | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|------|--------------|----------|-----------------------------|-----------------|--------------|--------|
| 08/18/2021 | Water  | M3484-01 | pH   | Cool 4 deg C | MINO01   | D11                         | O22-OUTFALL     | 08/18/2021   | 9040C  |

Date/Time 08-19-21 13:50  
Raw Sample Received by: JP (OC)  
Raw Sample Relinquished by: JP (OC)

Date/Time 08-19-21 13:50  
Raw Sample Received by: JP (OC)  
Raw Sample Relinquished by: JP (OC)