



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 8/20/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:45
In Date: 08/19/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:23
Out Date: 08/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115937

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3478-01	ETGI369	1	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3485-01	EO-02-081921	2	1.12	9.85	8.83	88.3	
M3485-02	EO-02-081921-EPH-2	3	1.16	9.68	8.59	87.2	
M3485-02D	EO-02-081921-EPH-2DUP	4	1.18	9.7	8.6	87.1	
M3489-01	1-BMENT-F	5	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3489-03	1-BMENT-W	6	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3489-05	1-EXT-A	7	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3489-07	1-EXT-B	8	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3489-09	2-GSLAB	9	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3489-11	2-EXT-A	10	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3489-13	2-EXT-B	11	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3489-15	1-GROWS	12	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3489-18	2-GROWS	13	1.00	2.00	2.00	100.0	CONCRETE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$