



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 08/20/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 08/19/2021 16:20 DATE & TIME OUT1: 08/20/2021 08:00OVEN IN TEMP2: 103 °COVEN OUT TEMP2: 103 °CDATE & TIME IN2: 08/20/2021 08:30 DATE & TIME OUT2: 08/20/2021 10:00Run Number: LB115939BalanceID: Wc Sc-05OvenID: Wc Oven-01

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB115939BL	LB115939BL	85.6371	85.6371	100	85.6372	85.6372	85.6372	0.0001	1
2	M3481-01	20211007-DI-WATER	75.5432	75.5432	100	75.5433	75.5433	75.5433	0.0001	1
3	M3481-01DUP	20211007-DI-WATERDUP	68.8104	68.8104	100	68.8105	68.8105	68.8105	0.0001	1

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

VB 115939

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ts m3492 WorkList ID : 151610 Department : Wet-Chemistry Date : 08-19-2021 15:52:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/25/2021	Water	M3481-01	TS	Cool 4 deg C	AQUA04	E11	20211007-DI-WATER	08/18/2021	SM2540 B

Date/Time 08.19.21 16:00
Raw Sample Received by: JR (WC)
Raw Sample Relinquished by: CR SM

Date/Time 08.19.21 17:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JR (WC)