



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115944

Slope : 98.2

BalanceID: Wc Sc-04

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/19/2021	17:35
2	CAL2	1	Water	NA	NA	21.2	7.00	08/19/2021	17:36
3	CAL3	1	Water	NA	NA	21.2	10.02	08/19/2021	17:37
4	ICV	1	Water	NA	NA	21.2	7.00	08/19/2021	17:40
5	CCV1	1	Water	NA	NA	21.3	2.01	08/19/2021	17:41
6	M3489-01	1	Solid	20.02	20	24.0	11.25	08/20/2021	17:50
7	M3489-03	1	Solid	20.03	20	24.4	9.50	08/19/2021	17:51
8	M3489-05	1	Solid	20.05	20	24.8	11.30	08/19/2021	17:55
9	M3489-07	1	Solid	20.02	20	24.3	10.35	08/19/2021	18:00
10	M3489-09	1	Solid	20.03	20	24.5	9.86	08/19/2021	18:02
11	M3489-11	1	Solid	20.04	20	24.4	9.48	08/19/2021	18:05
12	M3489-13	1	Solid	20.04	20	24.3	9.95	08/19/2021	18:06
13	M3489-15	1	Solid	20.03	20	24.3	10.97	08/19/2021	18:10
14	M3489-15DUP	1	Solid	20.04	20	24.4	10.98	08/19/2021	18:11
15	M3489-18	1	Solid	20.02	20	23.8	9.99	08/19/2021	18:12
16	CCV2	1	Water	NA	NA	21.2	12.02	08/19/2021	18:15

VB115944

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3489 WorkList ID : 151616 Department : Wet-Chemistry Date : 08-19-2021 16:41:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/19/2021	Solid	M3489-01	pH	Cool 4 deg C	PSEG03	E11	1-BMENT-F	08/19/2021	9045D
08/19/2021	Solid	M3489-03	pH	Cool 4 deg C	PSEG03	E11	1-BMENT-W	08/19/2021	9045D
08/19/2021	Solid	M3489-05	pH	Cool 4 deg C	PSEG03	E11	1-EXT-A	08/19/2021	9045D
08/19/2021	Solid	M3489-07	pH	Cool 4 deg C	PSEG03	E11	1-EXT-B	08/19/2021	9045D
08/19/2021	Solid	M3489-09	pH	Cool 4 deg C	PSEG03	E11	2-GSLAB	08/19/2021	9045D
08/19/2021	Solid	M3489-11	pH	Cool 4 deg C	PSEG03	E11	2-EXT-A	08/19/2021	9045D
08/19/2021	Solid	M3489-13	pH	Cool 4 deg C	PSEG03	E11	2-EXT-B	08/19/2021	9045D
08/19/2021	Solid	M3489-15	pH	Cool 4 deg C	PSEG03	E11	1-GROWS	08/19/2021	9045D
08/19/2021	Solid	M3489-18	pH	Cool 4 deg C	PSEG03	E11	2-GROWS	08/19/2021	9045D

08-19-21 16:43

Date/Time
Raw Sample Received by: JG (GSC)
Raw Sample Relinquished by: JG (GSC)Date/Time 08-19-21
Raw Sample Received by: JG (GSC)
Raw Sample Relinquished by: JG (GSC)