

Analytical Summary ReportAnalysis Method: SM4500 Cl GANALYST: apatelParameter: Residual ChlorineSUPERVISOR REVIEW BY: ketankumarRun Number: LB115950

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP92632
Residual chlorine ICV-LCS, 0.4PPM	WP92631
Chlorine Calibration std, 0.1ppm	WP92626
Chlorine Calibration std, 0.2ppm	WP92627
Chlorine Calibration std, 0.8ppm	WP92629
Chlorine Calibration std, 0.0ppm	WP92625
Chlorine Calibration std, 1.6ppm	WP92630
Residual Chlorine LOQ 0.1ppm	WP92633
Residual Chlorine Calibration and CCV std, 0	WP92628
Total Chlorine Powder Pillows	W2759
Total Chlorine Powder Pillows	W2759

Intercept: 0.0014Slope: 0.9488Regression: 0.999994

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	0.00		08/19/2021	11:52
2	CAL2	0.1	1	0.000	0.100	0.100	0.10	-4	08/19/2021	11:57
3	CAL3	0.2	1	0.000	0.190	0.190	0.20	0.5	08/19/2021	12:05
4	CAL4	0.4	1	0.000	0.380	0.380	0.40	0.3	08/19/2021	12:11
5	CAL5	0.8	1	0.000	0.760	0.760	0.80	0	08/19/2021	12:18
6	CAL6	1.6	1	0.000	1.520	1.520	1.60	-0.1	08/19/2021	12:25



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: apatel

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB115950

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3900	0.3900	0.4100	08/19/2021	12:32
2	ICB				1	0.0000	0.0000	0.0000	-0.0010	08/19/2021	12:38
3	CCV1			0.4	1	0.0000	0.3800	0.3800	0.3990	08/19/2021	12:45
4	CCB1				1	0.0000	0.0000	0.0000	-0.0010	08/19/2021	12:52
5	LB115950BL	50	50		1	0.0000	0.0100	0.0100	0.0090	08/19/2021	12:59
6	LB115950BS	50	50	0.4	1	0.0000	0.3800	0.3800	0.3990	08/19/2021	13:05
7	M2940-07	50	50		1	0.0000	0.0500	0.0500	0.0510	08/19/2021	13:11
8	M2940-08	50	50		1	0.0000	0.1000	0.1000	0.1040	08/19/2021	13:16
9	M3481-01	50	50		1	0.0000	0.0100	0.0100	0.0090	08/19/2021	13:22
10	M3481-01DUP	50	50		1	0.0000	0.0200	0.0200	0.0200	08/19/2021	13:28
11	CCV2			0.4	1	0.0000	0.3900	0.3900	0.4100	08/19/2021	13:35
12	CCB2				1	0.0000	0.0000	0.0000	-0.0010	08/19/2021	13:41

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Residual Chlorine-08192 WorkList ID : 151619 Department : Wet-Chemistry Date : 08-19-2021 14:09:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/18/2021	Water	M3481-01	Residual Chlorine	Cool 4 deg C	AQUA04	E11	20211007-DI-WATER	08/18/2021	SM4500 Cl G

Date/Time 8/19/21 11:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 8/19/21 12:45
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]