


LB115952

7470A

INSTRUMENT ID: CV



Linear

A= 0.0000e+000

B= 2.4796e-004

C= -6.2693e-003

Rho= 0.9998598

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%CD
0.0	0.000	-0.004	-0.004	9	0.000	9					-
0.2	0.200	0.209	0.009	869	0.0 %	869					4
2.5	2.500	2.433	-0.067	9837	0.0 %	9837					-3
5.0	5.000	5.047	0.047	20379	0.0 %	20379					1
7.5	7.500	7.597	0.097	30664	0.0 %	30664					1
10.0	10.000	9.918	-0.082	40025	0.0 %	40025					-1

LB115952 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 20 Aug 2021 08:49:28					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	9	-	0.00007470A	PPB		20 Aug 2021 08:51:19	S	Std
0.2 - 1	50.2	869	-	0.20007470A	PPB		20 Aug 2021 08:53:36	S	Std
2.5 - 1	52.5	9837	-	2.50007470A	PPB		20 Aug 2021 08:55:52	S	Std
5.0 - 1	55	20379	-	5.00007470A	PPB		20 Aug 2021 08:58:09	S	Std
7.5 - 1	57.5	30664	-	7.50007470A	PPB		20 Aug 2021 09:00:25	S	Std
10.0 - 1	510	40025	-	10.00007470A	PPB		20 Aug 2021 09:05:20	S	Std
ICV63 - 1	ICV63	15397	3.8115	-7470A	PPB		20 Aug 2021 09:08:12	U	SMPL
ICB63 - 1	ICB63	-89	-0.0283	-7470A	PPB		20 Aug 2021 09:10:27	U	SMPL
CCV14 - 1	CCV14	19978	4.9474	-7470A	PPB		20 Aug 2021 09:12:42	U	SMPL
CCB14 - 1	CCB14	-48	-0.0182	-7470A	PPB		20 Aug 2021 09:14:57	U	SMPL
CRA - 1	CRA	863	0.2077	-7470A	PPB		20 Aug 2021 09:17:12	U	SMPL
HighStd - 1	HighStd	40351	9.9990	-7470A	PPB		20 Aug 2021 09:19:27	U	SMPL
ChkStd - 1	ChkStd	27875	6.9055	-7470A	PPB		20 Aug 2021 09:24:54	U	SMPL
PB138572BL - 1	PBW	-110	-0.0335	-7470A	PPB		20 Aug 2021 09:27:08	U	SMPL
PB138572BS - 1	LCSW	16749	4.1467	-7470A	PPB		20 Aug 2021 09:29:25	U	SMPL
M3434-34 - 1	SWCS-15-0002	132	0.0265	-7470A	PPB		20 Aug 2021 09:31:41	U	SMPL
M3434-36 - 1	SWCS-15-0205	72	0.0116	-7470A	PPB		20 Aug 2021 09:33:57	U	SMPL
M3434-38 - 1	SWCS-15-0510	-11	-0.0090	-7470A	PPB		20 Aug 2021 09:36:13	U	SMPL
M3434-40 - 1	SWCS-15-1015	-10	-0.0087	-7470A	PPB		20 Aug 2021 09:38:31	U	SMPL
M3435-02 - 1	SWCS-17-0003	27	0.0004	-7470A	PPB		20 Aug 2021 09:40:46	U	SMPL
CCV15 - 1	CCV15	19988	4.9499	-7470A	PPB		20 Aug 2021 09:43:01	U	SMPL
CCB15 - 1	CCB15	-54	-0.0197	-7470A	PPB		20 Aug 2021 09:45:16	U	SMPL
M3435-04 - 1	SWCS-17-0306	47	0.0054	-7470A	PPB		20 Aug 2021 09:47:30	U	SMPL
M3435-06 - 1	SWCS-21-0002	52	0.0066	-7470A	PPB		20 Aug 2021 09:49:46	U	SMPL
M3435-08 - 1	SWCS-21-0204	29	0.0009	-7470A	PPB		20 Aug 2021 09:52:01	U	SMPL
M3435-10 - 1	SWCS-22-0306	-15	-0.0100	-7470A	PPB		20 Aug 2021 09:54:17	U	SMPL
M3435-12 - 1	2021-08-13-FD-04	15	-0.0025	-7470A	PPB		20 Aug 2021 09:56:33	U	SMPL
M3435-14 - 1	SWCS-24-0003	20	-0.0013	-7470A	PPB		20 Aug 2021 09:58:49	U	SMPL
M3435-16 - 1	SWCS-24-0306	18	-0.0018	-7470A	PPB		20 Aug 2021 10:01:06	U	SMPL
M3435-16DUP - 1	SWCS-24-0306DUP	10	-0.0038	-7470A	PPB		20 Aug 2021 10:03:22	U	SMPL
M3435-19 - 1	M3435-16MS	14673	3.6320	-7470A	PPB		20 Aug 2021 10:05:39	U	SMPL
M3435-20 - 1	M3435-16MSD	15175	3.7564	-7470A	PPB		20 Aug 2021 10:07:55	U	SMPL
CCV16 - 1	CCV16	19880	4.9231	-7470A	PPB		20 Aug 2021 10:10:10	U	SMPL
CCB16 - 1	CCB16	-37	-0.0154	-7470A	PPB		20 Aug 2021 10:12:25	U	SMPL
M3435-22 - 1	SWCS-26-0003	42	0.0041	-7470A	PPB		20 Aug 2021 10:14:41	U	SMPL
M3435-24 - 1	SWCS-26-0306	8	-0.0043	-7470A	PPB		20 Aug 2021 10:16:56	U	SMPL
M3435-26 - 1	SWCS-28-0003	16	-0.0023	-7470A	PPB		20 Aug 2021 10:19:11	U	SMPL
M3435-28 - 1	SWCS-28-0306	25	-0.0001	-7470A	PPB		20 Aug 2021 10:21:27	U	SMPL
M3435-30 - 1	SWCS-22-0003	5	-0.0050	-7470A	PPB		20 Aug 2021 10:23:42	U	SMPL
PB138573BL - 1	PBW	39	0.0034	-7470A	PPB		20 Aug 2021 10:25:59	U	SMPL
PB138573BS - 1	LCSW	15758	3.9010	-7470A	PPB		20 Aug 2021 10:28:15	U	SMPL
M3468-01 - 1	SAMPLE-11-(647.00)	-40	-0.0162	-7470A	PPB		20 Aug 2021 10:30:32	U	SMPL
M3468-02 - 1	SAMPLE-10-(410.75)	7	-0.0045	-7470A	PPB		20 Aug 2021 10:32:49	U	SMPL
M3468-03 - 1	SAMPLE-9-(322.00)	24	-0.0003	-7470A	PPB		20 Aug 2021 10:35:05	U	SMPL
CCV17 - 1	CCV17	19899	4.9278	-7470A	PPB		20 Aug 2021 10:37:21	U	SMPL
CCB17 - 1	CCB17	-67	-0.0229	-7470A	PPB		20 Aug 2021 10:39:37	U	SMPL
M3468-04 - 1	SAMPLE-8-(309.00)	7	-0.0045	-7470A	PPB		20 Aug 2021 10:41:52	U	SMPL
M3468-05 - 1	SAMPLE-7-(291.00)	28	0.0007	-7470A	PPB		20 Aug 2021 10:44:07	U	SMPL
M3468-06 - 1	SAMPLE-1-(150.00)	20	-0.0013	-7470A	PPB		20 Aug 2021 10:46:23	U	SMPL
M3468-07 - 1	SAMPLE-2-(161.00)	33	0.0019	-7470A	PPB		20 Aug 2021 10:48:39	U	SMPL
M3468-08 - 1	SAMPLE-3-(171.00)	24	-0.0003	-7470A	PPB		20 Aug 2021 10:50:55	U	SMPL
M3468-09 - 1	SAMPLE-4-(177.00)	43	0.0044	-7470A	PPB		20 Aug 2021 10:53:11	U	SMPL
M3468-10 - 1	SAMPLE-5-(184.50)	27	0.0004	-7470A	PPB		20 Aug 2021 10:55:28	U	SMPL
M3468-11 - 1	SAMPLE-6-(203.30)	22	-0.0008	-7470A	PPB		20 Aug 2021 10:57:44	U	SMPL
M3468-11DUP - 1	SAMPLE-6-(203.30)DUP	25	-0.0001	-7470A	PPB		20 Aug 2021 11:00:02	U	SMPL
M3468-11MS - 1	SAMPLE-6-(203.30)MS	13463	3.3320	-7470A	PPB		20 Aug 2021 11:02:17	U	SMPL
CCV18 - 1	CCV18	19347	4.7909	-7470A	PPB		20 Aug 2021 11:04:33	U	SMPL
CCB18 - 1	CCB18	-73	-0.0244	-7470A	PPB		20 Aug 2021 11:06:48	U	SMPL
M3468-11MSD - 1	SAMPLE-6-(203.30)MSD	15533	3.8452	-7470A	PPB		20 Aug 2021 11:09:04	U	SMPL
PB138574BL - 1	PBW	-54	-0.0197	-7470A	PPB		20 Aug 2021 11:11:19	U	SMPL
PB138574BS - 1	LCSW	15785	3.9077	-7470A	PPB		20 Aug 2021 11:18:21	U	SMPL
M3446-02 - 1	RW6B-GW-20210816	-80	-0.0261	-7470A	PPB		20 Aug 2021 11:20:37	U	SMPL
M3463-41 - 1	2021-08-17-EB	-28	-0.0132	-7470A	PPB		20 Aug 2021 11:22:53	U	SMPL
M3475-01 - 1	N5868-N34145	2	-0.0058	-7470A	PPB		20 Aug 2021 11:25:09	U	SMPL
M3475-01DUP - 1	N5868-N34145DUP	20	-0.0013	-7470A	PPB		20 Aug 2021 11:27:25	U	SMPL
M3475-01MS - 1	N5868-N34145MS	15150	3.7503	-7470A	PPB		20 Aug 2021 11:32:31	U	SMPL
M3475-01MSD - 1	N5868-N34145MSD	15538	3.8465	-7470A	PPB		20 Aug 2021 11:34:48	U	SMPL
CCV19 - 1	CCV19	19771	4.8961	-7470A	PPB		20 Aug 2021 11:37:03	U	SMPL
CCB19 - 1	CCB19	-83	-0.0268	-7470A	PPB		20 Aug 2021 11:39:19	U	SMPL
PB138572TB - 1	PB138544TB	-9	-0.0085	-7470A	PPB		20 Aug 2021 11:41:35	U	SMPL
PB138573TB - 1	PB138556TB	1	-0.0060	-7470A	PPB		20 Aug 2021 11:43:51	U	SMPL
M3435-16LX5 - 1		-13	-0.0095	-7470A	PPB		20 Aug 2021 11:46:08	U	SMPL
M3435-16A - 1		15636	3.8708	-7470A	PPB		20 Aug 2021 11:48:24	U	SMPL
M3468-11LX5 - 1		-81	-0.0264	-7470A	PPB		20 Aug 2021 11:50:40	U	SMPL
M3468-11A - 1		15799	3.9112	-7470A	PPB		20 Aug 2021 11:52:56	U	SMPL
M3475-01LX5 - 1		-84	-0.0271	-7470A	PPB		20 Aug 2021 11:55:13	U	SMPL
M3475-01A - 1		17893	4.4304	-7470A	PPB		20 Aug 2021 11:57:29	U	SMPL
CCV20 - 1	CCV20	19621	4.8589	-7470A	PPB		20 Aug 2021 11:59:46	U	SMPL
CCB20 - 1	CCB20	-72	-0.0241	-7470A	PPB		20 Aug 2021 12:02:02	U	SMPL