


LB115960

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.8878e-004

C= -3.2122e-002

Rho= 0.9996191

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.032	-0.032	0	0.000	0					-
0.2	0.200	0.190	-0.010	1176	0.0 %	1176					-5
2.5	2.500	2.429	-0.071	13035	0.0 %	13035					-3
5.0	5.000	5.147	0.147	27432	0.0 %	27432					3
7.5	7.500	7.614	0.114	40501	0.0 %	40501					2
10.0	10.000	9.853	-0.147	52363	0.0 %	52363					2

-1

LB
8/20/91

LB115960 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 20 Aug 2021 12:17:06				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	0	-	0.0000	7471B	PPB	20 Aug 2021 12:18:59	S Std
0.2 - 1	50.2	1176	-	0.2000	7471B	PPB	20 Aug 2021 12:21:15	S Std
2.5 - 1	52.5	13035	-	2.5000	7471B	PPB	20 Aug 2021 12:23:32	S Std
5.0 - 1	55	27432	-	5.0000	7471B	PPB	20 Aug 2021 12:25:49	S Std
7.5 - 1	57.5	40501	-	7.5000	7471B	PPB	20 Aug 2021 12:28:05	S Std
10.0 - 1	510	52363	-	10.0000	7471B	PPB	20 Aug 2021 12:30:21	S Std
ICV64 - 1	ICV64	20273	3.7951	-	7471B	PPB	20 Aug 2021 12:33:03	U SMPL
ICB64 - 1	ICB64	-22	-0.0363	-	7471B	PPB	20 Aug 2021 12:37:56	U SMPL
CCV21 - 1	CCV21	27526	5.1643	-	7471B	PPB	20 Aug 2021 12:40:11	U SMPL
CCB21 - 1	CCB21	-64	-0.0442	-	7471B	PPB	20 Aug 2021 12:42:26	U SMPL
CRA - 1	CRA	1166	0.1880	-	7471B	PPB	20 Aug 2021 12:44:42	U SMPL
HighStd - 1	HighStd	52682	9.9134	-	7471B	PPB	20 Aug 2021 12:46:57	U SMPL
ChkStd - 1	ChkStd	38471	7.2306	-	7471B	PPB	20 Aug 2021 12:49:17	U SMPL
PB138589BL - 1	PBS	26	-0.0272	-	7471B	PPB	20 Aug 2021 12:54:12	U SMPL
PB138589BS - 1	LCSS	21151	3.9608	-	7471B	PPB	20 Aug 2021 12:56:27	U SMPL
M3463-01 - 1	SWCS-20-0003	371	0.0379	-	7471B	PPB	20 Aug 2021 12:58:43	U SMPL
M3463-01DUP - 1	SWCS-20-0003DUP	395	0.0424	-	7471B	PPB	20 Aug 2021 13:03:39	U SMPL
M3463-03 - 1	M3463-01MS	19786	3.7032	-	7471B	PPB	20 Aug 2021 13:08:36	U SMPL
M3463-05 - 1	M3463-01MSD	19913	3.7271	-	7471B	PPB	20 Aug 2021 13:10:53	U SMPL
M3463-07 - 1	SWCS-20-0306	229	0.0111	-	7471B	PPB	20 Aug 2021 13:13:08	U SMPL
CCV22 - 1	CCV22	26374	4.9469	-	7471B	PPB	20 Aug 2021 13:15:24	U SMPL
CCB22 - 1	CCB22	76	-0.0178	-	7471B	PPB	20 Aug 2021 13:20:09	U SMPL
M3463-09 - 1	M3463-07MS	19752	3.6967	-	7471B	PPB	20 Aug 2021 13:22:23	U SMPL
M3463-11 - 1	M3463-07MSD	19555	3.6595	-	7471B	PPB	20 Aug 2021 13:24:38	U SMPL
M3463-13 - 1	SWCS-16-0002	2845	0.5050	-	7471B	PPB	20 Aug 2021 13:26:53	U SMPL
M3463-15 - 1	SWCS-16-0204	5823	1.0672	-	7471B	PPB	20 Aug 2021 13:29:09	U SMPL
M3463-17 - 1	SWCS-18-0002	8666	1.6039	-	7471B	PPB	20 Aug 2021 13:31:24	U SMPL
M3463-19 - 1	SWCS-18-0204	2153	0.3743	-	7471B	PPB	20 Aug 2021 13:33:41	U SMPL
M3463-21 - 1	SWCS-19-0002	1339	0.2207	-	7471B	PPB	20 Aug 2021 13:35:56	U SMPL
M3463-23 - 1	SWCS-19-0204	320	0.0283	-	7471B	PPB	20 Aug 2021 13:38:13	U SMPL
M3463-25 - 1	SWCS-23-0002	130	-0.0076	-	7471B	PPB	20 Aug 2021 13:40:30	U SMPL
M3463-27 - 1	SWCS-23-0204	733	0.1063	-	7471B	PPB	20 Aug 2021 13:42:46	U SMPL
CCV23 - 1	CCV23	26103	4.8957	-	7471B	PPB	20 Aug 2021 13:45:01	U SMPL
CCB23 - 1	CCB23	-27	-0.0372	-	7471B	PPB	20 Aug 2021 13:47:17	U SMPL
M3463-29 - 1	SWCS-25-0002	1638	0.2771	-	7471B	PPB	20 Aug 2021 13:49:32	U SMPL
M3463-31 - 1	SWCS-25-0204	392	0.0419	-	7471B	PPB	20 Aug 2021 13:51:48	U SMPL
M3463-33 - 1	SWCS-27-0002	363	0.0364	-	7471B	PPB	20 Aug 2021 13:54:03	U SMPL
M3463-35 - 1	SWCS-27-0204	94	-0.0144	-	7471B	PPB	20 Aug 2021 13:56:19	U SMPL
M3463-37 - 1	SWCS-29-0002	505	0.0632	-	7471B	PPB	20 Aug 2021 13:58:35	U SMPL
M3463-39 - 1	SWCS-29-0204	492	0.0608	-	7471B	PPB	20 Aug 2021 14:00:51	U SMPL
M3463-43 - 1	2021-08-17-FD-05	219	0.0092	-	7471B	PPB	20 Aug 2021 14:03:07	U SMPL
M3485-01 - 1	EO-02-081921	3284	0.5878	-	7471B	PPB	20 Aug 2021 14:05:24	U SMPL
M3463-01LX5 - 1		23	-0.0278	-	7471B	PPB	20 Aug 2021 14:07:42	U SMPL
M3463-01A - 1		20910	3.9153	-	7471B	PPB	20 Aug 2021 14:09:58	U SMPL
CCV24 - 1	CCV24	26006	4.8774	-	7471B	PPB	20 Aug 2021 14:12:13	U SMPL
CCB24 - 1	CCB24	51	-0.0225	-	7471B	PPB	20 Aug 2021 14:14:29	U SMPL