



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115970

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/20/2021	15:10
2	M3475-01		1	10	100.00	>212.0	>212.0	08/20/2021	15:40
3	M3475-01DUP		1	10	100.00	>212.0	>212.0	08/20/2021	16:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB115970

WorkList Name : FP81921 WorkList ID : 151623 Department : Wet-Chemistry Date : 08-19-2021 17:58:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/23/2021	Water	M3475-01	K Flash Point	Cool 4 deg C	PSEG03	E11	N5868-N34145	08/18/2021	1010B

Date/Time 08/20/21 14:45
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: Wah (DL)

Date/Time 08/20/21 16:10
 Raw Sample Received by: MS-DL
 Raw Sample Relinquished by: KS(wc)