


LB115971

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5496e-004

C= -4.9015e-002

Rho= 0.9996827

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.009	0.009	226	0.000	226					-
0.2	0.200	0.224	0.024	1070	0.0 %	1070					12
2.5	2.500	2.380	-0.120	9527	0.0 %	9527					-5
5.0	5.000	5.053	0.053	20012	0.0 %	20012					1
7.5	7.500	7.646	0.146	30182	0.0 %	30182					2
10.0	10.000	9.888	-0.112	38974	0.0 %	38974					-1

LB115971 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Aug 2021 15:49:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	226	-	0.0000	7470A	PPB	20 Aug 2021 15:49:16	S	Std
0.2 - 1	50.2	1070	-	0.2000	7470A	PPB	20 Aug 2021 15:51:33	S	Std
2.5 - 1	52.5	9527	-	2.5000	7470A	PPB	20 Aug 2021 15:53:50	S	Std
5.0 - 1	55	20012	-	5.0000	7470A	PPB	20 Aug 2021 15:56:07	S	Std
7.5 - 1	57.5	30182	-	7.5000	7470A	PPB	20 Aug 2021 15:58:23	S	Std
10.0 - 1	510	38974	-	10.0000	7470A	PPB	20 Aug 2021 16:03:11	S	Std
ICV65 - 1	ICV65	14735	3.7079	-	7470A	PPB	20 Aug 2021 16:05:53	U	SMPL
ICB65 - 1	ICB65	-73	-0.0676	-	7470A	PPB	20 Aug 2021 16:08:08	U	SMPL
CCV25 - 1	CCV25	19880	5.0197	-	7470A	PPB	20 Aug 2021 16:10:23	U	SMPL
CCB25 - 1	CCB25	-70	-0.0669	-	7470A	PPB	20 Aug 2021 16:12:38	U	SMPL
CRA - 1	CRA	1056	0.2202	-	7470A	PPB	20 Aug 2021 16:14:53	U	SMPL
HighStd - 1	HighStd	38381	9.7368	-	7470A	PPB	20 Aug 2021 16:17:10	U	SMPL
ChkStd - 1	ChkStd	27780	7.0339	-	7470A	PPB	20 Aug 2021 16:19:25	U	SMPL
PB138624BL - 1	PBW	-91	-0.0722	-	7470A	PPB	20 Aug 2021 16:21:42	U	SMPL
PB138624BS - 1	LCSW	15342	3.8626	-	7470A	PPB	20 Aug 2021 16:23:57	U	SMPL
M3361-20 - 1	41944	-21	-0.0544	-	7470A	PPB	20 Aug 2021 16:26:14	U	SMPL
M3361-21 - 1	41945	40	-0.0388	-	7470A	PPB	20 Aug 2021 16:28:30	U	SMPL
M3361-22 - 1	41946	27	-0.0421	-	7470A	PPB	20 Aug 2021 16:30:47	U	SMPL
M3361-23 - 1	41947	39	-0.0391	-	7470A	PPB	20 Aug 2021 16:33:05	U	SMPL
M3420-01 - 1	JONES-RD-STOCK-PILE	20	-0.0439	-	7470A	PPB	20 Aug 2021 16:35:20	U	SMPL
CCV26 - 1	CCV26	19672	4.9666	-	7470A	PPB	20 Aug 2021 16:37:35	U	SMPL
CCB26 - 1	CCB26	-68	-0.0664	-	7470A	PPB	20 Aug 2021 16:39:50	U	SMPL
M3489-16 - 1	1-GROWS	89	-0.0263	-	7470A	PPB	20 Aug 2021 16:42:05	U	SMPL
M3489-19 - 1	2-GROWS	31	-0.0411	-	7470A	PPB	20 Aug 2021 16:44:21	U	SMPL
M3489-19DUP - 1	2-GROWSDUP	38	-0.0393	-	7470A	PPB	20 Aug 2021 16:46:36	U	SMPL
M3489-19MS - 1	2-GROWSMS	14291	3.5947	-	7470A	PPB	20 Aug 2021 16:48:53	U	SMPL
M3489-19MSD - 1	2-GROWSMSD	13954	3.5088	-	7470A	PPB	20 Aug 2021 16:51:09	U	SMPL
PB138624TB - 1	PB138585TB	-33	-0.0574	-	7470A	PPB	20 Aug 2021 16:53:25	U	SMPL
M3489-19LX5 - 1		-21	-0.0544	-	7470A	PPB	20 Aug 2021 16:55:42	U	SMPL
M3489-19A - 1		14521	3.6533	-	7470A	PPB	20 Aug 2021 16:57:59	U	SMPL
CCV27 - 1	CCV27	19464	4.9136	-	7470A	PPB	20 Aug 2021 17:00:16	U	SMPL
CCB27 - 1	CCB27	-65	-0.0656	-	7470A	PPB	20 Aug 2021 17:02:31	U	SMPL