



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115973

Slope : 98.4

BalanceID: Wc Sc-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/20/2021	16:10
2	CAL2	1	Water	NA	NA	21.2	7.00	08/20/2021	16:11
3	CAL3	1	Water	NA	NA	21.2	10.03	08/20/2021	16:12
4	ICV	1	Water	NA	NA	21.2	7.00	08/20/2021	16:15
5	CCV1	1	Water	NA	NA	21.3	2.01	08/20/2021	16:19
6	M3502-02	1	Solid	20.03	20	23.9	5.95	08/20/2021	16:20
7	M3502-02DUP	1	Solid	20.04	20	23.8	5.96	08/20/2021	16:21
8	M3502-04	1	Solid	20.02	20	24.0	5.56	08/20/2021	16:24
9	M3505-02	1	Solid	20.03	20	24.1	9.43	08/20/2021	16:28
10	CCV2	1	Water	NA	NA	21.2	12.02	08/20/2021	16:30

VB 115973

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3502 WorkList ID : 151653 Department : Wet-Chemistry Date : 08-20-2021 15:53:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/19/2021	Solid	M3502-02	Corrosivity	Cool 4 deg C	NOBI03	D11	POST-MAEC-01	08/19/2021	9045D
08/19/2021	Solid	M3502-04	Corrosivity	Cool 4 deg C	NOBI03	D11	POST-MAEC-02	08/19/2021	9045D
08/20/2021	Solid	M3505-02	Corrosivity	Cool 4 deg C	WALS01	E12	DRILL-CUTTING	08/20/2021	9045D

Date/Time 08/20/21 1600
 Raw Sample Received by: JB FLOCC
 Raw Sample Relinquished by: JB FLOCC

Date/Time 08/20/21 17135
 Raw Sample Received by: JB FLOCC
 Raw Sample Relinquished by: JB FLOCC