



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115974

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/20/2021	16:10
2	CAL2	1	Water	NA	NA	21.2	7.00	08/20/2021	16:11
3	CAL3	1	Water	NA	NA	21.2	10.03	08/20/2021	16:22
4	ICV	1	Water	NA	NA	21.2	7.00	08/20/2021	16:15
5	CCV1	1	Water	NA	NA	21.3	2.01	08/20/2021	16:19
6	M3506-01	1	Water	NA	NA	20.8	8.01	08/20/2021	16:35
7	M3506-01DUP	1	Water	NA	NA	20.8	8.02	08/20/2021	16:36
8	M3512-01	1	Water	NA	NA	22.9	6.68	08/20/2021	16:40
9	CCV2	1	Water	NA	NA	21.3	12.02	08/20/2021	16:42

VB115974

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3512 WorkList ID : 151652 Department : Wet-Chemistry Date : 08-20-2021 15:53:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/20/2021	Water	M3506-01	pH	Cool 4 deg C	PSEG03	E12	N29305	08/20/2021	9040C
08/19/2021	Water	M3512-01	pH	Cool 4 deg C	STER01	E12	POOL-WATER	08/19/2021	9040C

Date/Time 08/20/21 16:00
 Raw Sample Received by: JB
 Raw Sample Relinquished by: A.C. Smith

Date/Time 08/20/21 18:15
 Raw Sample Received by: J.C. Smith
 Raw Sample Relinquished by: JP (WOC)