

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB115992

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2843
IODINE SOLUTION .025N 1L	W2763
Starch Solution, 4L	W2703

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB138670BL		1	5.00	50	2.00	0.00	1.92	1.92	0.08	0.00	0.00	08/23/2021	14:05
2	M3489-01		1	5.05	50	2.00	1.92	3.80	1.88	0.12	0.04	3.17	08/23/2021	14:08
3	M3489-01DUP		1	5.04	50	2.00	3.80	5.70	1.90	0.10	0.02	1.59	08/23/2021	14:12
4	M3489-03		1	5.05	50	2.00	5.70	7.56	1.86	0.14	0.06	4.75	08/23/2021	14:15
5	M3489-05		1	5.04	50	2.00	0.00	1.84	1.84	0.16	0.08	6.35	08/23/2021	14:18
6	M3489-07		1	5.03	50	2.00	1.84	3.70	1.86	0.14	0.06	4.77	08/23/2021	14:20
7	M3489-09		1	5.06	50	2.00	3.70	5.52	1.82	0.18	0.10	7.91	08/23/2021	14:22
8	M3489-11		1	5.04	50	2.00	5.52	7.36	1.84	0.16	0.08	6.35	08/23/2021	14:25
9	M3489-13		1	5.03	50	2.00	0.00	1.88	1.88	0.12	0.04	3.18	08/23/2021	14:28
10	M3489-16		1	5.04	50	2.00	1.88	3.72	1.84	0.16	0.08	6.35	08/23/2021	14:30
11	M3489-19		1	5.05	50	2.00	3.72	5.58	1.86	0.14	0.06	4.75	08/23/2021	14:32
12	M3502-02		1	5.03	50	2.00	5.58	7.40	1.82	0.18	0.10	7.95	08/23/2021	14:35

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB115992

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2843
IODINE SOLUTION .025N 1L	W2763
Starch Solution, 4L	W2703

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
13	M3502-04		1	5.05	50	2.00	0.00	1.86	1.86	0.14	0.06	4.75	08/23/2021	14:38
14	M3505-02		1	5.04	50	2.00	1.86	3.70	1.84	0.16	0.08	6.35	08/23/2021	14:40

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume