



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115997

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	08/23/2021	16:20
2	M3506-01		1	10	100.00	>212.0	>212.0	08/23/2021	16:45
3	M3506-01DUP		1	10	100.00	>212.0	>212.0	08/23/2021	17:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 115997

WorkList Name : fp82321 WorkList ID : 151691 Department : Wet-Chemistry Date : 08-23-2021 14:19:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/27/2021	Water	M3506-01	☒ Flash Point	Cool 4 deg C	PSEG03	E12	N29305	08/20/2021	1010B

Date/Time 08/23/21 15:55
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ CEXK-1067

Date/Time 08/23/21 17:15
 Raw Sample Received by: RJ CEXK-1067
 Raw Sample Relinquished by: KS (WC)