

LB116002


7470A

INSTRUMENT ID: CX1

Linear

 μ Abs.:

38698



A= 0.0000e+000

B= 2.5869e-004

C= -4.2203e-002

Rho= 0.9999774

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/011
0.0	0.000	-0.037	-0.037	20	0.000	20					6
0.2	0.200	0.212	0.012	983	0.0 %	983					1
2.5	2.500	2.517	0.017	9892	0.0 %	9892					0
5.0	5.000	5.012	0.012	19539	0.0 %	19539					0
7.5	7.500	7.527	0.027	29262	0.0 %	29262					0
10.0	10.000	9.968	-0.032	38698	0.0 %	38698					0

LB116002 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 24 Aug 2021 09:23:17				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	20	-	0.0000	7470A	PPB	24 Aug 2021 09:23:58	S Std
0.2 - 1	50.2	983	-	0.2000	7470A	PPB	24 Aug 2021 09:31:13	S Std
2.5 - 1	32.5	9892	-	2.5000	7470A	PPB	24 Aug 2021 09:33:30	S Std
5.0 - 1	55	19539	-	5.0000	7470A	PPB	24 Aug 2021 09:35:47	S Std
7.5 - 1	57.5	29262	-	7.5000	7470A	PPB	24 Aug 2021 09:44:14	S Std
10.0 - 1	510	38698	-	10.0000	7470A	PPB	24 Aug 2021 09:49:03	S Std
ICV66 - 1	ICV66	15160	3.8795	-	7470A	PPB	24 Aug 2021 09:52:05	U SMPL
ICB66 - 1	ICB66	-24	-0.0484	-	7470A	PPB	24 Aug 2021 09:54:20	U SMPL
CCV29 - 1	CCV29	19772	5.0726	-	7470A	PPB	24 Aug 2021 09:56:34	U SMPL
CCB29 - 1	CCB29	35	-0.0331	-	7470A	PPB	24 Aug 2021 09:58:49	U SMPL
CRA - 1	CRA	986	0.2129	-	7470A	PPB	24 Aug 2021 10:01:04	U SMPL
HighStd - 1	HighStd	38242	9.8505	-	7470A	PPB	24 Aug 2021 10:03:19	U SMPL
ChkStd - 1	ChkStd	28084	7.2228	-	7470A	PPB	24 Aug 2021 10:05:35	U SMPL
PB138623BL - 1	PBW	26	-0.0355	-	7470A	PPB	24 Aug 2021 10:07:51	U SMPL
PB138623BS - 1	LCSW	13748	3.5142	-	7470A	PPB	24 Aug 2021 10:10:07	U SMPL
M3463-02 - 1	SWCS-20-0003	51	-0.0290	-	7470A	PPB	24 Aug 2021 10:12:23	U SMPL
M3463-02DUP - 1	SWCS-20-0003DUP	109	-0.0140	-	7470A	PPB	24 Aug 2021 10:14:40	U SMPL
M3463-04 - 1	M3463-02MS	14105	3.6066	-	7470A	PPB	24 Aug 2021 10:16:56	U SMPL
M3463-06 - 1	M3463-02MSD	14345	3.6687	-	7470A	PPB	24 Aug 2021 10:19:13	U SMPL
M3463-08 - 1	SWCS-20-0306	135	-0.0073	-	7470A	PPB	24 Aug 2021 10:21:28	U SMPL
CCV30 - 1	CCV30	19859	5.0951	-	7470A	PPB	24 Aug 2021 10:23:44	U SMPL
CCB30 - 1	CCB30	35	-0.0331	-	7470A	PPB	24 Aug 2021 10:25:59	U SMPL
M3463-10 - 1	M3463-08MS	15315	3.9196	-	7470A	PPB	24 Aug 2021 10:28:15	U SMPL
M3463-12 - 1	M3463-08MSD	15009	3.8404	-	7470A	PPB	24 Aug 2021 10:30:30	U SMPL
M3471-01 - 1	SWCS-16-0002	424	0.0675	-	7470A	PPB	24 Aug 2021 10:32:46	U SMPL
M3471-02 - 1	SWCS-16-0204	318	0.0401	-	7470A	PPB	24 Aug 2021 10:35:01	U SMPL
M3471-03 - 1	SWCS-18-0002	149	-0.0037	-	7470A	PPB	24 Aug 2021 10:37:17	U SMPL
M3471-04 - 1	SWCS-18-0204	160	-0.0008	-	7470A	PPB	24 Aug 2021 10:39:33	U SMPL
M3471-05 - 1	SWCS-19-0002	139	-0.0062	-	7470A	PPB	24 Aug 2021 10:41:50	U SMPL
M3471-06 - 1	SWCS-19-0204	165	0.0005	-	7470A	PPB	24 Aug 2021 10:44:06	U SMPL
M3471-07 - 1	SWCS-23-0002	166	0.0007	-	7470A	PPB	24 Aug 2021 10:46:23	U SMPL
M3471-08 - 1	SWCS-23-0204	179	0.0041	-	7470A	PPB	24 Aug 2021 10:48:39	U SMPL
CCV31 - 1	CCV31	19588	5.0250	-	7470A	PPB	24 Aug 2021 10:50:54	U SMPL
CCB31 - 1	CCB31	54	-0.0282	-	7470A	PPB	24 Aug 2021 10:53:09	U SMPL
M3471-09 - 1	SWCS-25-0002	82	-0.0210	-	7470A	PPB	24 Aug 2021 10:55:25	U SMPL
M3471-10 - 1	SWCS-25-0204	292	0.0333	-	7470A	PPB	24 Aug 2021 10:57:40	U SMPL
M3471-11 - 1	SWCS-27-0002	438	0.0711	-	7470A	PPB	24 Aug 2021 10:59:56	U SMPL
M3471-12 - 1	SWCS-27-0204	480	0.0820	-	7470A	PPB	24 Aug 2021 11:02:12	U SMPL
M3471-13 - 1	SWCS-29-0002	443	0.0724	-	7470A	PPB	24 Aug 2021 11:04:27	U SMPL
M3471-14 - 1	SWCS-29-0204	383	0.0569	-	7470A	PPB	24 Aug 2021 11:06:44	U SMPL
M3471-16 - 1	2021-08-17-FD-05	-11	-0.0450	-	7470A	PPB	24 Aug 2021 11:09:00	U SMPL
PB138623TB - 1	PB138581TB	108	-0.0143	-	7470A	PPB	24 Aug 2021 11:11:17	U SMPL
M3463-02LX5 - 1		-6	-0.0438	-	7470A	PPB	24 Aug 2021 11:13:34	U SMPL
M3463-02A - 1		14776	3.7802	-	7470A	PPB	24 Aug 2021 11:17:50	U SMPL
CCV32 - 1	CCV32	19408	4.9784	-	7470A	PPB	24 Aug 2021 11:20:05	U SMPL
CCB32 - 1	CCB32	52	-0.0288	-	7470A	PPB	24 Aug 2021 11:22:20	U SMPL