


LB116004

SFAM01.1

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.3528e-004 slope

C= -6.8233e-002 y-intercept

Rho= 0.9999346

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.010	0.010	331	0.000	331					=
0.05	0.050					573					17
0.20	0.200	0.234	0.034	1285	0.0 %	1285					-2
2.50	2.500	2.461	-0.039	10752	0.0 %	10752					0
5.00	5.000	5.008	0.008	21577	0.0 %	21577					-1
7.50	7.500	7.431	-0.069	31873	0.0 %	31873					1
10.0	10.000	10.056	0.056	43029	0.0 %	43029					

LB116004
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	331	-		0 SFAM01.1	PPB	8/24/2021 12:00	Std
	0.2 S01	1285	-		0.2 SFAM01.1	PPB	8/24/2021 12:02	Std
	2.5 S02	10752	-		2.5 SFAM01.1	PPB	8/24/2021 12:04	Std
	5 S03	21577	-		5 SFAM01.1	PPB	8/24/2021 12:07	Std
	7.5 S04	31873	-		7.5 SFAM01.1	PPB	8/24/2021 12:09	Std
	10 S05	43029	-		10 SFAM01.1	PPB	8/24/2021 12:11	Std
ICV060	ICV060	17232	3.9861 -		SFAM01.1	PPB	8/24/2021 12:14	SMPL
ICB060	ICB060	241	-0.0115 -		SFAM01.1	PPB	8/24/2021 12:16	SMPL
CCV033	CCV033	21141	4.9058 -		SFAM01.1	PPB	8/24/2021 12:18	SMPL
CCB033	CCB033	247	-0.0101 -		SFAM01.1	PPB	8/24/2021 12:21	SMPL
PB138686BL	PBW686	452	0.0381 -		SFAM01.1	PPB	8/24/2021 12:23	SMPL
PB138686TB	LEB686	376	0.0202 -		SFAM01.1	PPB	8/24/2021 12:25	SMPL
M3488-01	MC0AB7	520	0.0541 -		SFAM01.1	PPB	8/24/2021 12:27	SMPL
M3488-02	MC0AB8	464	0.0409 -		SFAM01.1	PPB	8/24/2021 12:30	SMPL
M3488-03	MC0AB8D	466	0.0414 -		SFAM01.1	PPB	8/24/2021 12:36	SMPL
M3488-04	MC0AB8S	5104	1.1326 -		SFAM01.1	PPB	8/24/2021 12:38	SMPL
CCV034	CCV034	20711	4.8046 -		SFAM01.1	PPB	8/24/2021 12:40	SMPL
CCB034	CCB034	231	-0.0139 -		SFAM01.1	PPB	8/24/2021 12:43	SMPL