


LB116008

7471B

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.039	0.039	628	0.000	628					-
0.2	0.200	0.208	0.008	1463	0.0 %	1463					4
2.5	2.500	2.429	-0.071	12408	0.0 %	12408					-3
5.0	5.000	5.007	0.007	25111	0.0 %	25111					0
7.5	7.500	7.512	0.012	37455	0.0 %	37455					0
10.0	10.000	10.005	0.005	49740	0.0 %	49740					0

LB116008 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Aug 2021 13:58:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	628	-	0.0000	7471B	PPB	24 Aug 2021 14:13:07	S	Std
0.2 - 1	50.2	1463	-	0.2000	7471B	PPB	24 Aug 2021 14:15:24	S	Std
2.5 - 1	52.5	12408	-	2.5000	7471B	PPB	24 Aug 2021 14:17:41	S	Std
5.0 - 1	55	25111	-	5.0000	7471B	PPB	24 Aug 2021 14:19:58	S	Std
7.5 - 1	57.5	37455	-	7.5000	7471B	PPB	24 Aug 2021 14:22:15	S	Std
10.0 - 1	59	49740	-	10.0000	7471B	PPB	24 Aug 2021 14:27:02	S	Std
ICV68 - 1	ICV68	20235	4.0175	-	7471B	PPB	24 Aug 2021 14:29:48	U	SMPL
ICB68 - 1	ICB68	208	-0.0462	-	7471B	PPB	24 Aug 2021 14:32:02	U	SMPL
CCV35 - 1	CCV35	25022	4.9889	-	7471B	PPB	24 Aug 2021 14:34:17	U	SMPL
CCB35 - 1	CCB35	206	-0.0466	-	7471B	PPB	24 Aug 2021 14:36:32	U	SMPL
CRA - 1	CRA	1459	0.2076	-	7471B	PPB	24 Aug 2021 14:38:48	U	SMPL
HighStd - 1	HighStd	50099	10.0774	-	7471B	PPB	24 Aug 2021 14:41:03	U	SMPL
ChkStd - 1	ChkStd	35213	7.0568	-	7471B	PPB	24 Aug 2021 14:43:19	U	SMPL
PB138683BL - 1	PBS	250	-0.0377	-	7471B	PPB	24 Aug 2021 14:45:36	U	SMPL
PB138683BS - 1	LCSS	18323	3.6296	-	7471B	PPB	24 Aug 2021 14:47:52	U	SMPL
M3502-01 - 1	POST-MAEC-01	2169	0.3517	-	7471B	PPB	24 Aug 2021 14:50:08	U	SMPL
M3502-03 - 1	POST-MAEC-02	3812	0.6851	-	7471B	PPB	24 Aug 2021 14:52:25	U	SMPL
M3505-01 - 1	DRILL-CUTTING	704	0.0544	-	7471B	PPB	24 Aug 2021 14:54:42	U	SMPL
M3507-01 - 1	130-A	1738	0.2642	-	7471B	PPB	24 Aug 2021 14:56:59	U	SMPL
M3507-02 - 1	130-B	1920	0.3012	-	7471B	PPB	24 Aug 2021 14:59:14	U	SMPL
CCV36 - 1	CCV36	24783	4.9404	-	7471B	PPB	24 Aug 2021 15:01:30	U	SMPL
CCB36 - 1	CCB36	212	-0.0454	-	7471B	PPB	24 Aug 2021 15:03:45	U	SMPL
M3510-01 - 1	SU-02-082021	3636	0.6494	-	7471B	PPB	24 Aug 2021 15:06:00	U	SMPL
M3510-03 - 1	SU-03-082021	2046	0.3267	-	7471B	PPB	24 Aug 2021 15:08:15	U	SMPL
M3513-01 - 1	TR-04-082121	2776	0.4749	-	7471B	PPB	24 Aug 2021 15:10:30	U	SMPL
M3514-01 - 1	NB-8-082021	1084	0.1315	-	7471B	PPB	24 Aug 2021 15:12:46	U	SMPL
M3521-01 - 1	124019	3326	0.5865	-	7471B	PPB	24 Aug 2021 15:15:01	U	SMPL
M3521-03 - 1	124027	2624	0.4440	-	7471B	PPB	24 Aug 2021 15:17:17	U	SMPL
M3521-05 - 1	124035	2673	0.4540	-	7471B	PPB	24 Aug 2021 15:19:33	U	SMPL
M3523-01 - 1	72-11998	1339	0.1833	-	7471B	PPB	24 Aug 2021 15:21:51	U	SMPL
M3523-01DUP - 1	72-11998DUP	1332	0.1819	-	7471B	PPB	24 Aug 2021 15:24:08	U	SMPL
M3523-01MS - 1	72-11998MS	19315	3.8309	-	7471B	PPB	24 Aug 2021 15:26:24	U	SMPL
CCV37 - 1	CCV37	24610	4.9053	-	7471B	PPB	24 Aug 2021 15:28:39	U	SMPL
CCB37 - 1	CCB37	221	-0.0436	-	7471B	PPB	24 Aug 2021 15:30:55	U	SMPL
M3523-01MSD - 1	72-11998MSD	21140	4.2012	-	7471B	PPB	24 Aug 2021 15:33:11	U	SMPL
M3523-01LX5 - 1		731	0.0599	-	7471B	PPB	24 Aug 2021 15:35:26	U	SMPL
M3523-01A - 1		20235	4.0175	-	7471B	PPB	24 Aug 2021 15:37:42	U	SMPL
CCV38 - 1	CCV38	23680	4.7166	-	7471B	PPB	24 Aug 2021 15:39:58	U	SMPL
CCB38 - 1	CCB38	207	-0.0464	-	7471B	PPB	24 Aug 2021 15:42:14	U	SMPL