



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116010

Slope : 98.7

BalanceID: WC SC-04

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.3	4.01	08/24/2021	14:37
2	CAL2	1	Water	NA	NA	22.3	7.00	08/24/2021	14:38
3	CAL3	1	Water	NA	NA	22.3	10.03	08/24/2021	14:39
4	ICV	1	Water	NA	NA	22.4	6.99	08/24/2021	14:40
5	CCV1	1	Water	NA	NA	22.4	2.01	08/24/2021	14:41
6	M3528-02	1	Solid	20.03	20	23.6	6.74	08/24/2021	15:10
7	M3528-02DUP	1	Solid	20.04	20	23.7	6.75	08/24/2021	15:11
8	CCV2	1	Water	NA	NA	22.4	12.02	08/24/2021	15:15

VB116010

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3527 WorkList ID : 151715 Department : Wet-Chemistry Date : 08-24-2021 14:30:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/24/2021	Solid	M3528-02	Corrosivity	Cool 4 deg C	PSEG03	E22	PT-3-082421	08/24/2021	9045D

Date/Time 08/24/21 14:32
Raw Sample Received by: AKS
Raw Sample Relinquished by: AKS

Date/Time 08/24/21 15:39
Raw Sample Received by: AKS
Raw Sample Relinquished by: JP (WDC)