



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116016

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	08/25/2021	09:25
2	M3529-07		1	10	100.00	>212.0	>212.0	08/25/2021	09:45
3	M3529-07DUP		1	10	100.00	>212.0	>212.0	08/25/2021	10:10
4	M3529-08		1	12	100.00	>212.0	>212.0	08/25/2021	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB116016

WorkList Name : fp82521 WorkList ID : 151730 Department : Wet-Chemistry Date : 08-25-2021 08:57:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/31/2021	Water	M3529-07	Flash Point	Cool 4 deg C	PSEG03	E21	13-14-15	08/24/2021	1010B
08/31/2021	Water	M3529-08	Flash Point	Cool 4 deg C	PSEG03	E21	16-17-87	08/24/2021	1010B

Date/Time 08/25/21 09:05
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ (E-L-1ab)

Date/Time 08/25/21 10:40
 Raw Sample Received by: RJ (E-L-1ab)
 Raw Sample Relinquished by: KS (WC)