

LB116019


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37654



A= 0.0000e+000

B= 2.6885e-004

C= 3.3787e-002

Rho= 0.9995713

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.049	0.049	57	0.000	57					-
0.2	0.200	0.222	0.022	699	0.0 %	699					11
2.5	2.500	2.497	-0.003	9163	0.0 %	9163					0
5.0	5.000	4.980	-0.020	18397	0.0 %	18397					0
7.5	7.500	7.295	-0.205	27008	0.0 %	27008					-3
10.0	10.000	10.157	0.157	37654	0.0 %	37654					2

LB116019 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 25 Aug 2021 10:24:34					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	7470A	PPB	25 Aug 2021 10:30:16	S	Std
0.2 - 1	50.2	699	-	0.2000	7470A	PPB	25 Aug 2021 10:32:33	S	Std
2.5 - 1	52.1	9163	-	2.5000	7470A	PPB	25 Aug 2021 10:34:50	S	Std
5.0 - 1	55	18397	-	5.0000	7470A	PPB	25 Aug 2021 10:37:07	S	Std
7.5 - 1	57.1	27008	-	7.5000	7470A	PPB	25 Aug 2021 10:39:23	S	Std
10.0 - 1	51.8	37654	-	10.0000	7470A	PPB	25 Aug 2021 10:46:41	S	Std
ICV69 - 1	ICV69	14510	3.9348		7470A	PPB	25 Aug 2021 10:49:27	U	SMPL
ICB69 - 1	ICB69	-67	0.0158		7470A	PPB	25 Aug 2021 10:51:41	U	SMPL
CCV39 - 1	CCV39	18104	4.9011		7470A	PPB	25 Aug 2021 10:53:56	U	SMPL
CCB39 - 1	CCB39	-48	0.0209		7470A	PPB	25 Aug 2021 10:56:11	U	SMPL
CRA - 1	CRA	689	0.2190		7470A	PPB	25 Aug 2021 10:58:26	U	SMPL
HighStd - 1	HighStd	37187	10.0316		7470A	PPB	25 Aug 2021 11:00:41	U	SMPL
ChkStd - 1	ChkStd	25096	6.7809		7470A	PPB	25 Aug 2021 11:02:56	U	SMPL
PB138698BL - 1	PBW	-66	0.0160		7470A	PPB	25 Aug 2021 11:05:12	U	SMPL
PB138698BS - 1	LCSW	14989	4.0636		7470A	PPB	25 Aug 2021 11:07:29	U	SMPL
M3502-02 - 1	POST-MAEC-01	-9	0.0314		7470A	PPB	25 Aug 2021 11:09:45	U	SMPL
M3502-04 - 1	POST-MAEC-02	40	0.0445		7470A	PPB	25 Aug 2021 11:12:01	U	SMPL
M3505-02 - 1	DRILL-CUTTING	29	0.0416		7470A	PPB	25 Aug 2021 11:14:18	U	SMPL
M3521-02 - 1	124019	44	0.0456		7470A	PPB	25 Aug 2021 11:16:35	U	SMPL
M3521-04 - 1	124027	33	0.0427		7470A	PPB	25 Aug 2021 11:18:50	U	SMPL
CCV40 - 1	CCV40	17801	4.8196		7470A	PPB	25 Aug 2021 11:21:05	U	SMPL
CCB40 - 1	CCB40	-47	0.0212		7470A	PPB	25 Aug 2021 11:23:21	U	SMPL
M3521-06 - 1	124035	30	0.0419		7470A	PPB	25 Aug 2021 11:25:36	U	SMPL
M3521-06DUP - 1	124035DUP	47	0.0464		7470A	PPB	25 Aug 2021 11:27:51	U	SMPL
M3521-06MS - 1	124035MS	12283	3.3361		7470A	PPB	25 Aug 2021 11:30:06	U	SMPL
M3521-06MSD - 1	124035MSD	12942	3.5133		7470A	PPB	25 Aug 2021 11:35:01	U	SMPL
PB138698TB - 1	PB138664TB	-23	0.0276		7470A	PPB	25 Aug 2021 11:37:16	U	SMPL
M3521-06A-1	06-LX5	38	0.0440		7470A	PPB	25 Aug 2021 11:39:32	U	SMPL
M3521-06A - 1		14332	3.8870		7470A	PPB	25 Aug 2021 11:41:48	U	SMPL
CCV41 - 1	CCV41	17969	4.8648		7470A	PPB	25 Aug 2021 11:44:05	U	SMPL
CCB41 - 1	CCB41	-34	0.0246		7470A	PPB	25 Aug 2021 11:46:22	U	SMPL